

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030623\
 Data File : VX034343.D
 Acq On : 06 Mar 2023 12:36
 Operator : JC/MD
 Sample : 01758-01DL 10X
 Misc : 1.09g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FMC-0081DL

Quant Time: Mar 07 00:36:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	279315	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	468782	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	420155	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	195744	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	168540	44.804	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.600%
35) Dibromofluoromethane	5.385	113	145073	47.530	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.060%
50) Toluene-d8	8.647	98	565871	51.637	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.280%
62) 4-Bromofluorobenzene	11.079	95	198543	47.366	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.740%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	5464	2.816	ug/l	99
39) Methylcyclohexane	7.379	83	24352	4.192	ug/l	96
52) Toluene	8.720	92	386460	47.679	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030623\
Data File : VX034343.D
Acq On : 06 Mar 2023 12:36
Operator : JC/MD
Sample : 01758-01DL 10X
Misc : 1.09g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FMC-0081DL

Quant Time: Mar 07 00:36:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

