

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030623\
 Data File : VX034348.D
 Acq On : 06 Mar 2023 14:36
 Operator : JC/MD
 Sample : 01821-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-2023-3-3

Quant Time: Mar 07 00:38:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	268750	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	449397	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	398351	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	176649	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	165783	45.804	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.600%
35) Dibromofluoromethane	5.385	113	140315	47.954	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.900%
50) Toluene-d8	8.647	98	536662	51.084	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.160%
62) 4-Bromofluorobenzene	11.079	95	185005	46.040	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.080%
Target Compounds						
20) Methylene Chloride	2.788	84	1682	0.500	ug/l	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030623\
Data File : VX034348.D
Acq On : 06 Mar 2023 14:36
Operator : JC/MD
Sample : 01821-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-2023-3-3

Quant Time: Mar 07 00:38:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

