

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030623\
 Data File : VX034353.D
 Acq On : 06 Mar 2023 16:35
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 07 00:40:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	277013	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	476599	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	422757	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	209909	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	179624	48.147	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.300%
35) Dibromofluoromethane	5.385	113	148227	47.767	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.540%
50) Toluene-d8	8.647	98	574876	51.599	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.200%
62) 4-Bromofluorobenzene	11.079	95	209678	49.202	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.400%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	1123	0.324	ug/l #	88
84) 1,2,4-Trimethylbenzene	11.756	105	6934	0.510	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030623\
Data File : VX034353.D
Acq On : 06 Mar 2023 16:35
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 07 00:40:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

