



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030623\
Data File : VX034335.D
Acq On : 06 Mar 2023 08:41
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Title : SW846 8260
Last Update : Fri Feb 17 12:33:35 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.0	20092	PASS
75	95	30	60	51.1	51464	PASS
95	95	100	100	100.0	100632	PASS
96	95	5	9	6.5	6561	PASS
173	174	0.00	2	0.8	671	PASS
174	95	50	100	78.4	78944	PASS
175	174	5	9	7.3	5767	PASS
176	174	95	101	97.6	77048	PASS
177	176	5	9	6.5	5010	PASS