

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
 Data File : VX029561.D
 Acq On : 18 Jun 2022 09:14
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Quant Time: Jun 18 10:34:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 10:29:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	264609	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	464548	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	416950	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	193345	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	326688	77.173	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 154.340%#	
35) Dibromofluoromethane	5.391	113	291494	76.139	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 152.280%#	
50) Toluene-d8	8.653	98	1071332	78.263	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 156.520%#	
62) 4-Bromofluorobenzene	11.085	95	405331	76.208	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 152.420%#	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	333067	82.800	ug/l	98
3) Chloromethane	1.288	50	318567	74.281	ug/l	100
4) Vinyl Chloride	1.374	62	320529	73.380	ug/l	98
5) Bromomethane	1.605	94	139207	60.818	ug/l	100
6) Chloroethane	1.685	64	127972	55.409	ug/l	100
7) Trichlorofluoromethane	1.880	101	439578	69.420	ug/l	100
8) Diethyl Ether	2.136	74	175527	70.753	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.325	101	273907	71.236	ug/l	98
10) Methyl Iodide	2.453	142	328388	78.128	ug/l	98
11) Tert butyl alcohol	2.983	59	107829	89.999	ug/l	95
12) 1,1-Dichloroethene	2.319	96	273757	73.436	ug/l	95
13) Acrolein	2.239	56	93133	217.681	ug/l	99
14) Allyl chloride	2.666	41	443461	74.003	ug/l	95
15) Acrylonitrile	3.068	53	840905	357.834	ug/l	99
16) Acetone	2.386	43	652398	326.902	ug/l	97
17) Carbon Disulfide	2.508	76	729160	78.936	ug/l	99
18) Methyl Acetate	2.709	43	368644	69.472	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	953922	74.163	ug/l	99
20) Methylene Chloride	2.788	84	305191	68.048	ug/l	97
21) trans-1,2-Dichloroethene	3.093	96	301279	71.908	ug/l	95
22) Diisopropyl ether	3.763	45	831296	69.287	ug/l	93
23) Vinyl Acetate	3.727	43	3607148	363.641	ug/l	100
24) 1,1-Dichloroethane	3.611	63	520620	71.720	ug/l	99
25) 2-Butanone	4.568	43	1100348	349.244	ug/l	99
26) 2,2-Dichloropropane	4.477	77	413804	78.772	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	350599	70.724	ug/l	94
28) Bromochloromethane	4.903	49	193611	68.080	ug/l	99
29) Tetrahydrofuran	5.013	42	727532	353.966	ug/l	96
30) Chloroform	5.099	83	540841	68.964	ug/l	99
31) Cyclohexane	5.470	56	463439	70.043	ug/l	96
32) 1,1,1-Trichloroethane	5.385	97	482387	71.433	ug/l	98
36) 1,1-Dichloropropene	5.696	75	403685	66.542	ug/l	99
37) Ethyl Acetate	4.721	43	424990	68.077	ug/l	99
38) Carbon Tetrachloride	5.684	117	416482	67.274	ug/l	97
39) Methylcyclohexane	7.385	83	511394	68.206	ug/l	95
40) Benzene	6.044	78	1206902	66.781	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
 Data File : VX029561.D
 Acq On : 18 Jun 2022 09:14
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Quant Time: Jun 18 10:34:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 10:29:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	224583	72.985	ug/l	97
42) 1,2-Dichloroethane	6.092	62	401563	64.258	ug/l	98
43) Isopropyl Acetate	6.348	43	669501	71.001	ug/l	100
44) Trichloroethene	7.129	130	339408	68.446	ug/l	97
45) 1,2-Dichloropropane	7.434	63	299170	67.482	ug/l	98
46) Dibromomethane	7.586	93	218288	67.877	ug/l	94
47) Bromodichloromethane	7.824	83	417707	68.854	ug/l	98
48) Methyl methacrylate	7.696	41	323718	68.519	ug/l	94
49) 1,4-Dioxane	7.683	88	162407	1346.509	ug/l	97
51) 4-Methyl-2-Pentanone	8.580	43	1988867	325.011	ug/l	100
52) Toluene	8.720	92	770486	66.593	ug/l	97
53) t-1,3-Dichloropropene	8.982	75	464493	78.701	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	501974	73.844	ug/l	97
55) 1,1,2-Trichloroethane	9.153	97	313295	66.917	ug/l	98
56) Ethyl methacrylate	9.122	69	501929	73.945	ug/l	97
57) 1,3-Dichloropropane	9.311	76	514985	65.227	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	1137927	344.649	ug/l	96
59) 2-Hexanone	9.439	43	1500982	323.286	ug/l	98
60) Dibromochloromethane	9.525	129	328268	72.629	ug/l	97
61) 1,2-Dibromoethane	9.610	107	337340	67.702	ug/l	98
64) Tetrachloroethene	9.275	164	293591	77.905	ug/l	94
65) Chlorobenzene	10.079	112	822422	75.466	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	308929	81.704	ug/l	98
67) Ethyl Benzene	10.195	91	1468259	76.962	ug/l	98
68) m/p-Xylenes	10.305	106	1141706	154.910	ug/l	99
69) o-Xylene	10.646	106	562767	78.503	ug/l	100
70) Styrene	10.659	104	953958	79.651	ug/l	99
71) Bromoform	10.805	173	225903	84.061	ug/l	99
73) Isopropylbenzene	10.963	105	1447852	96.947	ug/l	99
74) N-amyl acetate	10.848	43	525083	103.692	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.213	83	447694	89.127	ug/l	99
76) 1,2,3-Trichloropropane	11.244	75	509100	115.111	ug/l	89
77) Bromobenzene	11.201	156	338331	96.437	ug/l	96
78) n-propylbenzene	11.305	91	1651302	97.481	ug/l	99
79) 2-Chlorotoluene	11.366	91	993831	95.198	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	1229201	96.923	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.024	75	148403	111.660	ug/l	99
82) 4-Chlorotoluene	11.457	91	1134153	94.393	ug/l	99
83) tert-Butylbenzene	11.719	119	1191449	96.941	ug/l	95
84) 1,2,4-Trimethylbenzene	11.756	105	1217257	97.260	ug/l	99
85) sec-Butylbenzene	11.890	105	1488144	98.326	ug/l	99
86) p-Isopropyltoluene	12.012	119	1261835	99.454	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	632458	94.438	ug/l	98
88) 1,4-Dichlorobenzene	12.042	146	632548	92.782	ug/l	97
89) n-Butylbenzene	12.335	91	1090836	102.015	ug/l	99
90) Hexachloroethane	12.542	117	201715	100.409	ug/l	91
91) 1,2-Dichlorobenzene	12.335	146	610674	92.360	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.945	75	112180	106.891	ug/l	98
93) 1,2,4-Trichlorobenzene	13.591	180	407916	103.513	ug/l	98
94) Hexachlorobutadiene	13.725	225	157611	95.836	ug/l	94
95) Naphthalene	13.780	128	1486471	107.523	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	410100	104.622	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
Data File : VX029561.D
Acq On : 18 Jun 2022 09:14
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100


Quant Time: Jun 18 10:34:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 10:29:26 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 18 10:34:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 10:29:26 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:168 Resp: 264609
Ion Ratio Lower Upper
168 100
99 60.8 41.7 62.5



#2
Dichlorodifluoromethane
Concen: 82.800 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion: 85 Resp: 333067
Ion Ratio Lower Upper
85 100
87 33.6 16.2 48.6









#3
Chloromethane
Concen: 74.281 ug/l
RT: 1.288 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100


Tgt Ion: 50 Resp: 318567
Ion Ratio Lower Upper
50 100
52 32.9 26.4 39.6



#4
Vinyl Chloride
Concen: 73.380 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion: 62 Resp: 320529
Ion Ratio Lower Upper
62 100
64 33.1 25.5 38.3









#5
Bromomethane
Concen: 60.818 ug/l
RT: 1.605 min Scan# 81
Delta R.T. -0.006 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 94 Resp: 139207
Ion Ratio Lower Upper
94 100
96 94.0 75.3 112.9



#6
Chloroethane
Concen: 55.409 ug/l
RT: 1.685 min Scan# 98
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion: 64 Resp: 127972
Ion Ratio Lower Upper
64 100
66 31.6 25.3 37.9









#7
 Trichlorofluoromethane
 Concen: 69.420 ug/l
 RT: 1.880 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: VX029561.D
 Acq: 18 Jun 2022 09:14




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100


Tgt Ion:101 Resp: 439578
 Ion Ratio Lower Upper
 101 100
 103 63.4 50.5 75.7



#8
 Diethyl Ether
 Concen: 70.753 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. -0.000 min
 Lab File: VX029561.D
 Acq: 18 Jun 2022 09:14

Tgt Ion: 74 Resp: 175527
 Ion Ratio Lower Upper
 74 100
 45 83.6 42.0 126.0









#9
1,1,2-Trichlorotrifluoroethane
Concen: 71.236 ug/l
RT: 2.325 min Scan# 204
Delta R.T. -0.006 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:101 Resp: 273907
Ion Ratio Lower Upper
101 100
85 42.2 33.5 50.3
151 75.2 61.8 92.6



#10
Methyl Iodide
Concen: 78.128 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion:142 Resp: 328388
Ion Ratio Lower Upper
142 100
127 44.5 33.9 50.9
141 14.3 11.4 17.2









#11
Tert butyl alcohol
Concen: 89.999 ug/l
RT: 2.983 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 59 Resp: 107829
Ion Ratio Lower Upper
59 100
57 8.7 8.3 12.5



#12
1,1-Dichloroethene
Concen: 73.436 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion: 96 Resp: 273757
Ion Ratio Lower Upper
96 100
61 148.4 124.6 187.0
98 65.3 51.0 76.4





#13

Acrolein

Concen: 217.681 ug/l

RT: 2.239 min Scan# 189

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC100

Tgt Ion: 56 Resp: 93133

Ion Ratio Lower Upper

56 100

55 69.8 55.0 82.6



#14

Allyl chloride

Concen: 74.003 ug/l

RT: 2.666 min Scan# 259

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14




Tgt Ion: 41 Resp: 443461


Ion Ratio Lower Upper

41 100

39 59.8 50.4 75.6

76 36.9 32.7 49.1









#15
Acrylonitrile
Concen: 357.834 ug/l
RT: 3.068 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100


Tgt Ion: 53 Resp: 840905
Ion Ratio Lower Upper
53 100
52 81.3 65.7 98.5
51 35.5 28.4 42.6



#16
Acetone
Concen: 326.902 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion: 43 Resp: 652398
Ion Ratio Lower Upper
43 100
58 34.3 28.8 43.2





#17

Carbon Disulfide

Concen: 78.936 ug/l

RT: 2.508 min Scan# 21

Delta R.T. -0.006 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC100

Tgt Ion: 76 Resp: 729160

Ion Ratio Lower Upper

76 100

78 9.2 7.2 10.8



#18

Methyl Acetate

Concen: 69.472 ug/l

RT: 2.709 min Scan# 266

Delta R.T. -0.000 min

Lab File: VX029561.D




Acq: 18 Jun 2022 09:14


Tgt Ion: 43 Resp: 368644

Ion Ratio Lower Upper

43 100

74 27.3 22.8 34.2









#19
Methyl tert-butyl Ether
Concen: 74.163 ug/l
RT: 3.117 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 73 Resp: 953922
Ion Ratio Lower Upper
73 100
57 20.5 16.8 25.2



#20
Methylene Chloride
Concen: 68.048 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion: 84 Resp: 305191
Ion Ratio Lower Upper
84 100
49 111.6 85.5 128.3
51 33.7 26.6 39.8
86 64.2 51.5 77.3









#21
trans-1,2-Dichloroethene
Concen: 71.908 ug/l
RT: 3.093 min Scan# 312
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 96 Resp: 301279
Ion Ratio Lower Upper
96 100
61 126.2 105.4 158.2
98 65.7 50.0 75.0



#22
Diisopropyl ether
Concen: 69.287 ug/l
RT: 3.763 min Scan# 439
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion: 45 Resp: 831296
Ion Ratio Lower Upper
45 100
43 72.1 50.8 76.2
87 33.4 26.6 39.8
59 12.8 10.5 15.7









#23
 Vinyl Acetate
 Concen: 363.641 ug/l
 RT: 3.727 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX029561.D
 Acq: 18 Jun 2022 09:14




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100


Tgt Ion: 43 Resp: 3607148
 Ion Ratio Lower Upper
 43 100
 86 13.5 10.8 16.2



#24
 1,1-Dichloroethane
 Concen: 71.720 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. -0.000 min
 Lab File: VX029561.D
 Acq: 18 Jun 2022 09:14

Tgt Ion: 63 Resp: 520620
 Ion Ratio Lower Upper
 63 100
 98 7.7 3.9 11.7
 100 5.2 2.4 7.2









#25
2-Butanone
Concen: 349.244 ug/l
RT: 4.568 min Scan# 51
Delta R.T. 0.006 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 43 Resp: 1100348
Ion Ratio Lower Upper
43 100
72 29.9 24.4 36.6



#26
2,2-Dichloropropane
Concen: 78.772 ug/l
RT: 4.477 min Scan# 556
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion: 77 Resp: 413804
Ion Ratio Lower Upper
77 100
97 24.2 11.8 35.4





#27

cis-1,2-Dichloroethene

Concen: 70.724 ug/l

RT: 4.489 min Scan# 51

Delta R.T. -0.006 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion: 96 Resp: 350599

Ion Ratio Lower Upper

96 100

61 131.7 0.0 283.8

98 64.5 0.0 129.8



#28

Bromochloromethane

Concen: 68.080 ug/l

RT: 4.903 min Scan# 626

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14




Tgt Ion: 49 Resp: 193611


Ion Ratio Lower Upper

49 100

129 2.3 0.0 4.4

130 89.2 70.7 106.1





#29

Tetrahydrofuran

Concen: 353.966 ug/l

RT: 5.013 min Scan# 644

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100





Tgt Ion: 42 Resp: 727532

Ion Ratio Lower Upper

42 100

72 52.1 44.1 66.1

71 48.0 40.5 60.7



#30

Chloroform

Concen: 68.964 ug/l

RT: 5.099 min Scan# 658

Delta R.T. -0.000 min

Lab File: VX029561.D




Acq: 18 Jun 2022 09:14


Tgt Ion: 83 Resp: 540841

Ion Ratio Lower Upper

83 100


85 65.2 52.7 79.1








#31
Cyclohexane
Concen: 70.043 ug/l
RT: 5.470 min Scan# 719
Delta R.T. -0.006 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 56 Resp: 463439




Ion	Ratio	Lower	Upper
56	100		
69	31.9	26.7	40.1
84	92.0	77.2	115.8




#32
1,1,1-Trichloroethane
Concen: 71.433 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion: 97 Resp: 482387

Ion	Ratio	Lower	Upper
97	100		
99	65.1	51.8	77.6
61	39.4	34.0	51.0





#33

1,2-Dichloroethane-d4

Concen: 77.173 ug/l

RT: 5.958 min Scan# 799

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC100

Tgt Ion: 65 Resp: 326688

Ion Ratio Lower Upper

65 100

67 53.5 0.0 107.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14




Tgt Ion: 114 Resp: 464548


Ion Ratio Lower Upper

114 100

63 19.4 0.0 36.8

88 16.3 0.0 32.4









#35
 Dibromofluoromethane
 Concen: 76.139 ug/l
 RT: 5.391 min Scan# 706
 Delta R.T. -0.000 min
 Lab File: VX029561.D
 Acq: 18 Jun 2022 09:14




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100


Tgt Ion:113 Resp: 291494
 Ion Ratio Lower Upper
 113 100
 111 102.8 82.8 124.2
 192 19.1 16.9 25.3



#36
 1,1-Dichloropropene
 Concen: 66.542 ug/l
 RT: 5.696 min Scan# 756
 Delta R.T. -0.000 min
 Lab File: VX029561.D
 Acq: 18 Jun 2022 09:14

Tgt Ion: 75 Resp: 403685
 Ion Ratio Lower Upper
 75 100
 110 37.7 18.3 54.8
 77 31.0 24.8 37.2









#37
Ethyl Acetate
Concen: 68.077 ug/l
RT: 4.721 min Scan# 596
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 43 Resp: 424990
Ion Ratio Lower Upper
43 100
61 15.7 12.6 19.0
70 13.4 10.2 15.2



#38
Carbon Tetrachloride
Concen: 67.274 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion: 117 Resp: 416482
Ion Ratio Lower Upper
117 100
119 93.5 77.3 115.9
121 30.5 24.2 36.2





#39

Methylcyclohexane

Concen: 68.206 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion: 83 Resp: 511394

Ion Ratio Lower Upper

83 100

55 71.3 54.6 81.8

98 51.2 37.4 56.2



#40

Benzene

Concen: 66.781 ug/l

RT: 6.044 min Scan# 813

Delta R.T. -0.000 min

Lab File: VX029561.D




Acq: 18 Jun 2022 09:14


Tgt Ion: 78 Resp: 1206902

Ion Ratio Lower Upper

78 100

77 23.0 19.3 28.9









#41
Methacrylonitrile
Concen: 72.985 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 41 Resp: 224583
Ion Ratio Lower Upper
41 100
39 52.9 43.0 64.4
67 83.0 70.2 105.4
52 31.2 26.0 39.0



#42
1,2-Dichloroethane
Concen: 64.258 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion: 62 Resp: 401563
Ion Ratio Lower Upper
62 100
98 10.8 0.0 20.4





#43

Isopropyl Acetate

Concen: 71.001 ug/l

RT: 6.348 min Scan# 863

Delta R.T. 0.006 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion: 43 Resp: 669501

Ion Ratio Lower Upper

43 100

61 22.9 18.5 27.7

87 16.4 13.0 19.6



#44

Trichloroethene

Concen: 68.446 ug/l

RT: 7.129 min Scan# 991

Delta R.T. -0.000 min

Lab File: VX029561.D




Acq: 18 Jun 2022 09:14


Tgt Ion:130 Resp: 339408

Ion Ratio Lower Upper

130 100

95 97.6 0.0 188.8





#45

1,2-Dichloropropane

Concen: 67.482 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC100

Tgt Ion: 63 Resp: 299170

Ion Ratio Lower Upper

63 100

65 31.1 25.8 38.6



#46

Dibromomethane

Concen: 67.877 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14




Tgt Ion: 93 Resp: 218288


Ion Ratio Lower Upper

93 100

95 83.0 65.8 98.6

174 97.4 85.4 128.2





#47

Bromodichloromethane

Concen: 68.854 ug/l

RT: 7.824 min Scan# 1105

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion: 83 Resp: 417707

Ion Ratio Lower Upper

83 100

85 63.8 52.2 78.2

127 8.5 6.8 10.2



#48

Methyl methacrylate

Concen: 68.519 ug/l

RT: 7.696 min Scan# 1084

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14




Tgt Ion: 41 Resp: 323718


Ion Ratio Lower Upper

41 100

69 96.0 82.5 123.7

39 52.3 43.8 65.8





#49

1,4-Dioxane

Concen: 1346.509 ug/l

RT: 7.683 min Scan# 1082

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion: 88 Resp: 162407

Ion Ratio Lower Upper

88 100

43 26.1 21.9 32.9

58 64.7 53.7 80.5



#50

Toluene-d8

Concen: 78.263 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. -0.000 min

Lab File: VX029561.D




Acq: 18 Jun 2022 09:14


Tgt Ion: 98 Resp: 1071332

Ion Ratio Lower Upper

98 100

100 67.6 52.6 78.8





#51

4-Methyl-2-Pentanone

Concen: 325.011 ug/l

RT: 8.580 min Scan# 1229

Delta R.T. 0.006 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC100

Tgt Ion: 43 Resp: 1988867

Ion Ratio Lower Upper

43 100

58 44.9 36.1 54.1



#52

Toluene

Concen: 66.593 ug/l

RT: 8.720 min Scan# 1252

Delta R.T. -0.000 min

Lab File: VX029561.D




Acq: 18 Jun 2022 09:14


Tgt Ion: 92 Resp: 770486

Ion Ratio Lower Upper

92 100

91 168.0 137.5 206.3









#53
 t-1,3-Dichloropropene
 Concen: 78.701 ug/l
 RT: 8.982 min Scan# 1194
 Delta R.T. -0.000 min
 Lab File: VX029561.D
 Acq: 18 Jun 2022 09:14




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100


Tgt Ion: 75 Resp: 464493
 Ion Ratio Lower Upper
 75 100
 77 31.7 25.4 38.0



#54
 cis-1,3-Dichloropropene
 Concen: 73.844 ug/l
 RT: 8.366 min Scan# 1194
 Delta R.T. -0.000 min
 Lab File: VX029561.D
 Acq: 18 Jun 2022 09:14

Tgt Ion: 75 Resp: 501974
 Ion Ratio Lower Upper
 75 100
 77 31.3 24.8 37.2
 39 40.0 29.6 44.4









#55
1,1,2-Trichloroethane
Concen: 66.917 ug/l
RT: 9.153 min Scan# 1318
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100

Tgt Ion: 97 Resp: 313295




Ion	Ratio	Lower	Upper
97	100		
83	84.1	66.2	99.4
85	55.9	42.9	64.3
99	63.0	49.0	73.6




#56
Ethyl methacrylate
Concen: 73.945 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.006 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion: 69 Resp: 501929

Ion	Ratio	Lower	Upper
69	100		
41	57.6	43.8	65.6
39	28.9	22.2	33.2





#57

1,3-Dichloropropane

Concen: 65.227 ug/l

RT: 9.311 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC100

Tgt Ion: 76 Resp: 514985

Ion Ratio Lower Upper

76 100

78 32.8 25.8 38.8



#58

2-Chloroethyl Vinyl ether

Concen: 344.649 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX029561.D




Acq: 18 Jun 2022 09:14


Tgt Ion: 63 Resp: 1137927

Ion Ratio Lower Upper

63 100

106 32.3 23.9 35.9





#59

2-Hexanone

Concen: 323.286 ug/l

RT: 9.439 min Scan# 1370

Delta R.T. 0.006 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC100

Tgt Ion: 43 Resp: 1500982

Ion Ratio Lower Upper

43 100

58 62.3 31.9 95.7



#60

Dibromochloromethane

Concen: 72.629 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. -0.000 min

Lab File: VX029561.D




Acq: 18 Jun 2022 09:14


Tgt Ion: 129 Resp: 328268

Ion Ratio Lower Upper

129 100

127 77.7 40.3 120.8









#61
1,2-Dibromoethane
Concen: 67.702 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:107 Resp: 337340
Ion Ratio Lower Upper
107 100
109 94.9 74.4 111.6



#62
4-Bromofluorobenzene
Concen: 76.208 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion: 95 Resp: 405331
Ion Ratio Lower Upper
95 100
174 73.7 0.0 163.0
176 70.9 0.0 155.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100


Tgt Ion:117 Resp: 416950
Ion Ratio Lower Upper
117 100
82 55.1 41.8 62.8
119 31.7 24.8 37.2



#64
Tetrachloroethene
Concen: 77.905 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion:164 Resp: 293591
Ion Ratio Lower Upper
164 100
166 128.2 101.2 151.8
129 104.2 73.0 109.4
131 97.9 74.2 111.2









#65
Chlorobenzene
Concen: 75.466 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:112 Resp: 822422
Ion Ratio Lower Upper
112 100
114 33.1 25.8 38.8



#66
1,1,1,2-Tetrachloroethane
Concen: 81.704 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion:131 Resp: 308929
Ion Ratio Lower Upper
131 100
133 94.5 48.3 144.8
119 64.0 32.6 98.0









#67
Ethyl Benzene
Concen: 76.962 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 91 Resp: 1468259
Ion Ratio Lower Upper
91 100
106 31.7 24.6 36.8



#68
m/p-Xylenes
Concen: 154.910 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion: 106 Resp: 1141706
Ion Ratio Lower Upper
106 100
91 198.2 160.4 240.6





#69

o-Xylene

Concen: 78.503 ug/l

RT: 10.646 min Scan# 1568

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC100

Tgt Ion:106 Resp: 562767

Ion Ratio Lower Upper

106 100

91 212.1 106.1 318.1



#70

Styrene

Concen: 79.651 ug/l

RT: 10.659 min Scan# 1570

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14




Tgt Ion:104 Resp: 953958


Ion Ratio Lower Upper

104 100

78 49.4 38.6 57.8

103 54.0 43.8 65.8





#71

Bromoform

Concen: 84.061 ug/l

RT: 10.805 min Scan# 11

Delta R.T. 0.006 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion:173 Resp: 225903

Ion Ratio Lower Upper

173 100

175 48.6 24.1 72.2

254 0.1 0.1 0.1



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14




Tgt Ion:152 Resp: 193345


Ion Ratio Lower Upper

152 100

115 110.3 37.5 112.7

150 192.4 0.0 338.2









#73
Isopropylbenzene
Concen: 96.947 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:105 Resp: 1447852
Ion Ratio Lower Upper
105 100
120 26.7 13.5 40.5



#74
N-ethyl acetate
Concen: 103.692 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion: 43 Resp: 525083
Ion Ratio Lower Upper
43 100
70 51.1 42.5 63.7
55 26.0 23.1 34.7
61 27.7 22.8 34.2









#75
 1,1,2,2-Tetrachloroethane
 Concen: 89.127 ug/l
 RT: 11.213 min Scan# 1661
 Delta R.T. -0.000 min
 Lab File: VX029561.D
 Acq: 18 Jun 2022 09:14




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100


Tgt Ion: 83 Resp: 447694
 Ion Ratio Lower Upper
 83 100
 131 10.7 5.1 15.3
 85 64.8 32.6 97.7



#76
 1,2,3-Trichloropropane
 Concen: 115.111 ug/l
 RT: 11.244 min Scan# 1666
 Delta R.T. -0.000 min
 Lab File: VX029561.D
 Acq: 18 Jun 2022 09:14

Tgt Ion: 75 Resp: 509100
 Ion Ratio Lower Upper
 75 100
 77 32.2 19.4 58.2





#77

Bromobenzene

Concen: 96.437 ug/l

RT: 11.201 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100





Tgt Ion:156 Resp: 338331

Ion Ratio Lower Upper

156 100

77 162.0 78.5 235.3

158 95.8 50.2 150.7



#78

n-propylbenzene

Concen: 97.481 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. -0.000 min

Lab File: VX029561.D




Acq: 18 Jun 2022 09:14


Tgt Ion: 91 Resp: 1651302

Ion Ratio Lower Upper

91 100

120 23.6 11.6 34.8









#79
2-Chlorotoluene
Concen: 95.198 ug/l
RT: 11.366 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 91 Resp: 993831
Ion Ratio Lower Upper
91 100
126 33.9 17.3 51.9



#80
1,3,5-Trimethylbenzene
Concen: 96.923 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion: 105 Resp: 1229201
Ion Ratio Lower Upper
105 100
120 48.8 24.3 72.9









#81
trans-1,4-Dichloro-2-butene
Concen: 111.660 ug/l
RT: 11.024 min Scan# 1630
Delta R.T. 0.006 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 75 Resp: 148403
Ion Ratio Lower Upper
75 100
53 86.8 69.0 103.4
89 45.5 37.0 55.6



#82
4-Chlorotoluene
Concen: 94.393 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion: 91 Resp: 1134153
Ion Ratio Lower Upper
91 100
126 30.1 14.9 44.5





#83

tert-Butylbenzene

Concen: 96.941 ug/l

RT: 11.719 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100





Tgt Ion:119 Resp: 1191449

Ion Ratio Lower Upper

119 100

91 58.5 26.9 80.6

134 23.2 11.4 34.2



#84

1,2,4-Trimethylbenzene

Concen: 97.260 ug/l

RT: 11.756 min Scan# 1750

Delta R.T. -0.000 min

Lab File: VX029561.D




Acq: 18 Jun 2022 09:14


Tgt Ion:105 Resp: 1217257

Ion Ratio Lower Upper

105 100

120 44.8 22.6 67.8









#85
 sec-Butylbenzene
 Concen: 98.326 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX029561.D
 Acq: 18 Jun 2022 09:14




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100


Tgt Ion:105 Resp: 1488144
 Ion Ratio Lower Upper
 105 100
 134 20.7 10.2 30.4



#86
 p-Isopropyltoluene
 Concen: 99.454 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. -0.000 min
 Lab File: VX029561.D
 Acq: 18 Jun 2022 09:14

Tgt Ion:119 Resp: 1261835
 Ion Ratio Lower Upper
 119 100
 134 26.4 13.0 39.0
 91 24.2 11.3 33.9





#87

1,3-Dichlorobenzene

Concen: 94.438 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion:146 Resp: 632458

Ion Ratio Lower Upper

146 100

111 42.6 19.7 59.1

148 63.5 31.6 94.7



#88

1,4-Dichlorobenzene

Concen: 92.782 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14




Tgt Ion:146 Resp: 632548


Ion Ratio Lower Upper

146 100

111 41.4 19.2 57.6

148 63.4 32.3 96.9









#89
n-Butylbenzene
Concen: 102.015 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 91 Resp: 1090836
Ion Ratio Lower Upper
91 100
92 55.8 27.8 83.3
134 27.1 13.3 39.8



#90
Hexachloroethane
Concen: 100.409 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion: 117 Resp: 201715
Ion Ratio Lower Upper
117 100
201 70.8 39.5 118.5









#91
1,2-Dichlorobenzene
Concen: 92.360 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:146 Resp: 610674
Ion Ratio Lower Upper
146 100
111 44.2 21.0 63.0
148 64.5 32.6 97.7



#92
1,2-Dibromo-3-Chloropropane
Concen: 106.891 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX029561.D
Acq: 18 Jun 2022 09:14

Tgt Ion: 75 Resp: 112180
Ion Ratio Lower Upper
75 100
155 91.1 47.0 141.2
157 116.0 57.9 173.6





#93

1,2,4-Trichlorobenzene

Concen: 103.513 ug/l

RT: 13.591 min Scan# 2051

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100





Tgt Ion:180 Resp: 407916

Ion Ratio Lower Upper

180 100

182 95.3 47.7 143.1

145 33.6 15.2 45.5



#94

Hexachlorobutadiene

Concen: 95.836 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14




Tgt Ion:225 Resp: 157611


Ion Ratio Lower Upper

225 100

223 63.0 34.2 102.5

227 63.5 33.9 101.7





#95

Naphthalene

Concen: 107.523 ug/l

RT: 13.780 min Scan# 2081

Delta R.T. 0.006 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion:128 Resp: 1486471

Ion Ratio Lower Upper

128 100

127 13.0 10.2 15.4

129 11.2 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 104.622 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. -0.000 min

Lab File: VX029561.D

Acq: 18 Jun 2022 09:14

Tgt Ion:180 Resp: 410100

Ion Ratio Lower Upper

180 100

182 94.9 50.0 149.8

145 35.3 16.2 48.5

