

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
 Data File : VX029565.D
 Acq On : 18 Jun 2022 13:21
 Operator : JC/MD
 Sample : VX0618WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0618WBL01

Quant Time: Jun 20 01:32:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	272335	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	453650	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	391690	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	172223	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	152317	42.252	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.500%
35) Dibromofluoromethane	5.385	113	132729	44.896	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.800%
50) Toluene-d8	8.647	98	513447	47.541	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.080%
62) 4-Bromofluorobenzene	11.079	95	185974	45.662	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
Data File : VX029565.D
Acq On : 18 Jun 2022 13:21
Operator : JC/MD
Sample : VX0618WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0618WBL01

Quant Time: Jun 20 01:32:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

