

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
 Data File : VX029574.D
 Acq On : 18 Jun 2022 16:59
 Operator : JC/MD
 Sample : N3325-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-23

Quant Time: Jun 20 01:35:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	270838	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	454189	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	389547	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	175617	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	155542	43.385	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.780%
35) Dibromofluoromethane	5.391	113	135357	45.731	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.460%
50) Toluene-d8	8.653	98	515112	47.638	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.280%
62) 4-Bromofluorobenzene	11.079	95	190023	46.600	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.200%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	4347	3.042	ug/l	93
17) Carbon Disulfide	2.520	76	3122	0.439	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	8421	0.884	ug/l #	1
31) Cyclohexane	5.471	56	6757	1.404	ug/l	93
40) Benzene	6.044	78	66152	5.571	ug/l	98
68) m/p-Xylenes	10.305	106	2086	0.389	ug/l	100
73) Isopropylbenzene	10.963	105	18112	1.376	ug/l	100
78) n-propylbenzene	11.305	91	16257	1.096	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	5092	0.460	ug/l	95
95) Naphthalene	13.780	128	3587	0.291	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
Data File : VX029574.D
Acq On : 18 Jun 2022 16:59
Operator : JC/MD
Sample : N3325-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-23

Quant Time: Jun 20 01:35:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

