

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
 Data File : VX029576.D
 Acq On : 18 Jun 2022 17:46
 Operator : JC/MD
 Sample : N3325-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-25

Quant Time: Jun 20 01:35:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	270128	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	460228	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	395064	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	172854	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	158775	44.403	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.800%
35) Dibromofluoromethane	5.391	113	136139	45.391	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.780%
50) Toluene-d8	8.652	98	516492	47.139	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.280%
62) 4-Bromofluorobenzene	11.085	95	186789	45.206	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.420%
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	30740	21.568	ug/l	99
17) Carbon Disulfide	2.513	76	2874	0.405	ug/l #	80
25) 2-Butanone	4.586	43	2042	0.906	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
Data File : VX029576.D
Acq On : 18 Jun 2022 17:46
Operator : JC/MD
Sample : N3325-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-25

Quant Time: Jun 20 01:35:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

