

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
 Data File : VX029592.D
 Acq On : 19 Jun 2022 00:44
 Operator : JC/MD
 Sample : N3290-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-22

Quant Time: Jun 20 01:38:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	290045	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	480396	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	419239	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	179000	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	162968	42.446	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.900%
35) Dibromofluoromethane	5.391	113	144484	46.151	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.300%
50) Toluene-d8	8.653	98	548721	47.978	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.960%
62) 4-Bromofluorobenzene	11.079	95	197063	45.690	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.380%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	3480	2.274	ug/l	100
17) Carbon Disulfide	2.514	76	6485	0.851	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
Data File : VX029592.D
Acq On : 19 Jun 2022 00:44
Operator : JC/MD
Sample : N3290-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-22

Quant Time: Jun 20 01:38:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

