

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
 Data File : VX029593.D
 Acq On : 19 Jun 2022 01:07
 Operator : JC/MD
 Sample : N3290-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-24

Quant Time: Jun 20 01:39:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	290735	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	488890	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	425923	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	188378	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	166164	43.176	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	86.360%		
35) Dibromofluoromethane	5.391	113	145187	45.570	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.140%		
50) Toluene-d8	8.653	98	556048	47.774	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.540%		
62) 4-Bromofluorobenzene	11.085	95	203037	46.258	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.520%		
Target Compounds						
11) Tert butyl alcohol	2.983	59	32147	30.195	ug/l	100
17) Carbon Disulfide	2.514	76	2018	0.264	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
Data File : VX029593.D
Acq On : 19 Jun 2022 01:07
Operator : JC/MD
Sample : N3290-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-24

Quant Time: Jun 20 01:39:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

