

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
 Data File : VX029594.D
 Acq On : 19 Jun 2022 01:30
 Operator : JC/MD
 Sample : N3290-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-25

Quant Time: Jun 20 01:39:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	279771	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	472261	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	408618	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	181776	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	159699	43.122	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.240%
35) Dibromofluoromethane	5.391	113	141109	45.850	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.700%
50) Toluene-d8	8.653	98	533021	47.408	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.820%
62) 4-Bromofluorobenzene	11.079	95	193262	45.581	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.160%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.977	59	103855	124.022	ug/l	99
16) Acetone	2.386	43	4092	2.772	ug/l	95
19) Methyl tert-butyl Ether	3.117	73	11253	1.144	ug/l	94
22) Diisopropyl ether	3.763	45	10773	1.190	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
Data File : VX029594.D
Acq On : 19 Jun 2022 01:30
Operator : JC/MD
Sample : N3290-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-25

Quant Time: Jun 20 01:39:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

