

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
 Data File : VX029596.D
 Acq On : 19 Jun 2022 02:17
 Operator : JC/MD
 Sample : N3290-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-28D2

Quant Time: Jun 20 01:39:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	273804	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	465466	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	402651	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	178750	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	156807	43.264	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	86.520%		
35) Dibromofluoromethane	5.391	113	137973	45.485	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	90.980%		
50) Toluene-d8	8.653	98	526066	47.472	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	94.940%		
62) 4-Bromofluorobenzene	11.079	95	190856	45.671	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	91.340%		
Target Compounds						Qvalue
11) Tert butyl alcohol	2.971	59	579019	751.656	ug/l	99
16) Acetone	2.386	43	7834	5.423	ug/l	94
19) Methyl tert-butyl Ether	3.117	73	1159555	120.409	ug/l	99
22) Diisopropyl ether	3.763	45	22391	2.527	ug/l #	83
40) Benzene	6.050	78	7949	0.653	ug/l	96
67) Ethyl Benzene	10.195	91	4629	0.326	ug/l #	89
68) m/p-Xylenes	10.305	106	13553	2.445	ug/l	99
69) o-Xylene	10.646	106	1494	0.273	ug/l	89
78) n-propylbenzene	11.305	91	7513	0.498	ug/l	93
80) 1,3,5-Trimethylbenzene	11.451	105	2463	0.218	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	4892	0.435	ug/l	98
95) Naphthalene	13.780	128	14356	1.145	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
Data File : VX029596.D
Acq On : 19 Jun 2022 02:17
Operator : JC/MD
Sample : N3290-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-28D2

Quant Time: Jun 20 01:39:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

