

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
 Data File : VX029598.D
 Acq On : 19 Jun 2022 03:03
 Operator : JC/MD
 Sample : N3290-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-33

Quant Time: Jun 20 01:40:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	270723	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	454312	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	389388	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	166054	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	154884	43.220	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.440%
35) Dibromofluoromethane	5.391	113	135026	45.606	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.220%
50) Toluene-d8	8.653	98	518489	47.937	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.880%
62) 4-Bromofluorobenzene	11.079	95	183371	44.957	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.920%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.977	59	94900	116.580	ug/l	100
19) Methyl tert-butyl Ether	3.117	73	716034	75.199	ug/l	100
40) Benzene	6.050	78	12748	1.073	ug/l	97
68) m/p-Xylenes	10.299	106	3156	0.589	ug/l	95
73) Isopropylbenzene	10.963	105	3519	0.283	ug/l	100
78) n-propylbenzene	11.305	91	6068	0.433	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
Data File : VX029598.D
Acq On : 19 Jun 2022 03:03
Operator : JC/MD
Sample : N3290-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-33

Quant Time: Jun 20 01:40:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

