

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
 Data File : VX029603.D
 Acq On : 19 Jun 2022 04:59
 Operator : JC/MD
 Sample : N3290-22 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-3

Quant Time: Jun 20 01:41:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	270505	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	460107	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	402160	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	172994	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	158565	44.283	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	88.560%		
35) Dibromofluoromethane	5.385	113	135362	45.144	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	90.280%		
50) Toluene-d8	8.653	98	522400	47.691	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.380%		
62) 4-Bromofluorobenzene	11.079	95	188858	45.719	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	91.440%		
Target Compounds						
16) Acetone	2.386	43	2502	1.753	ug/l	88
52) Toluene	8.726	92	2203	0.286	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
Data File : VX029603.D
Acq On : 19 Jun 2022 04:59
Operator : JC/MD
Sample : N3290-22 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW-3

Quant Time: Jun 20 01:41:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

