

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
 Data File : VX029604.D
 Acq On : 19 Jun 2022 05:23
 Operator : JC/MD
 Sample : N3290-21 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-2

Quant Time: Jun 20 01:41:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	274133	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	460724	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	397445	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	169133	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	156923	43.244	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.480%
35) Dibromofluoromethane	5.391	113	136533	45.474	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.940%
50) Toluene-d8	8.653	98	520084	47.416	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.840%
62) 4-Bromofluorobenzene	11.079	95	183908	44.461	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.920%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	2672	1.847	ug/l #	86
40) Benzene	6.044	78	2703	0.224	ug/l	91
52) Toluene	8.726	92	2237	0.290	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
Data File : VX029604.D
Acq On : 19 Jun 2022 05:23
Operator : JC/MD
Sample : N3290-21 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW-2

Quant Time: Jun 20 01:41:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

