

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX061822\
 Data File : VX029586.D
 Acq On : 18 Jun 2022 21:39
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 20 01:37:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	50.000	50.535	-1.1	81	0.00
3 P	Chloromethane	50.000	48.982	2.0	87	0.00
4 C	Vinyl Chloride	50.000	51.063	-2.1#	86	0.00
5 T	Bromomethane	50.000	44.819	10.4	84	0.00
6 T	Chloroethane	50.000	54.219	-8.4	89	0.00
7 T	Trichlorofluoromethane	50.000	49.291	1.4	82	0.00
8 T	Diethyl Ether	50.000	57.856	-15.7	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	46.576	6.8	80	0.00
10 T	Methyl Iodide	50.000	26.051	47.9#	43	0.00
11 T	Tert butyl alcohol	250.000	287.547	-15.0	103	0.00
12 CM	1,1-Dichloroethene	50.000	52.130	-4.3#	88	0.00
13 T	Acrolein	250.000	263.580	-5.4	96	0.00
14 T	Allyl chloride	50.000	53.557	-7.1	90	0.00
15 T	Acrylonitrile	250.000	291.584	-16.6	98	0.00
16 T	Acetone	250.000	274.564	-9.8	100	0.00
17 T	Carbon Disulfide	50.000	50.122	-0.2	85	0.00
18 T	Methyl Acetate	50.000	56.701	-13.4	98	0.00
19 T	Methyl tert-butyl Ether	50.000	58.916	-17.8	99	0.00
20 T	Methylene Chloride	50.000	54.553	-9.1	95	0.00
21 T	trans-1,2-Dichloroethene	50.000	53.082	-6.2	89	0.00
22 T	Diisopropyl ether	50.000	57.119	-14.2	96	0.00
23 T	Vinyl Acetate	250.000	296.933	-18.8	97	0.00
24 P	1,1-Dichloroethane	50.000	54.341	-8.7	92	0.00
25 T	2-Butanone	250.000	291.304	-16.5	99	0.00
26 T	2,2-Dichloropropane	50.000	39.287	21.4	65	0.00
27 T	cis-1,2-Dichloroethene	50.000	54.279	-8.6	92	0.00
28 T	Bromochloromethane	50.000	56.238	-12.5	97	0.00
29 T	Tetrahydrofuran	250.000	288.951	-15.6	97	0.00
30 C	Chloroform	50.000	54.831	-9.7#	92	0.00
31 T	Cyclohexane	50.000	46.515	7.0	79	0.00
32 T	1,1,1-Trichloroethane	50.000	51.054	-2.1	85	0.00
33 S	1,2-Dichloroethane-d4	50.000	54.014	-8.0	93	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	90	0.00
35 S	Dibromofluoromethane	50.000	52.435	-4.9	92	0.00
36 T	1,1-Dichloropropene	50.000	47.422	5.2	83	0.00
37 T	Ethyl Acetate	50.000	57.115	-14.2	97	0.00
38 T	Carbon Tetrachloride	50.000	47.429	5.1	81	0.00
39 T	Methylcyclohexane	50.000	44.912	10.2	77	0.00
40 TM	Benzene	50.000	51.566	-3.1	90	0.00
41 T	Methacrylonitrile	50.000	58.192	-16.4	97	0.00
42 TM	1,2-Dichloroethane	50.000	53.522	-7.0	95	0.00
43 T	Isopropyl Acetate	50.000	57.124	-14.2	97	0.00
44 TM	Trichloroethene	50.000	47.824	4.4	85	0.00
45 C	1,2-Dichloropropane	50.000	53.638	-7.3#	95	0.00
46 T	Dibromomethane	50.000	54.510	-9.0	94	0.00
47 T	Bromodichloromethane	50.000	53.771	-7.5	92	0.00
48 T	Methyl methacrylate	50.000	56.394	-12.8	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
 Data File : VX029586.D
 Acq On : 18 Jun 2022 21:39
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 20 01:37:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1098.059	-9.8	97	0.00
50 S	Toluene-d8	50.000	50.164	-0.3	86	0.00
51 T	4-Methyl-2-Pentanone	250.000	283.428	-13.4	98	0.00
52 CM	Toluene	50.000	50.622	-1.2#	86	0.00
53 T	t-1,3-Dichloropropene	50.000	53.582	-7.2	86	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.801	-5.6	86	0.00
55 T	1,1,2-Trichloroethane	50.000	55.325	-10.7	95	0.00
56 T	Ethyl methacrylate	50.000	58.618	-17.2	97	0.00
57 T	1,3-Dichloropropane	50.000	54.347	-8.7	95	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	265.416	-6.2	89	0.00
59 T	2-Hexanone	250.000	289.375	-15.8	98	0.00
60 T	Dibromochloromethane	50.000	55.955	-11.9	93	0.00
61 T	1,2-Dibromoethane	50.000	55.821	-11.6	94	0.00
62 S	4-Bromofluorobenzene	50.000	50.550	-1.1	85	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	88	0.00
64 T	Tetrachloroethene	50.000	47.059	5.9	82	0.00
65 PM	Chlorobenzene	50.000	50.769	-1.5	87	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.359	-4.7	87	0.00
67 C	Ethyl Benzene	50.000	49.872	0.3#	85	0.00
68 T	m/p-Xylenes	100.000	99.174	0.8	85	0.00
69 T	o-Xylene	50.000	50.637	-1.3	87	0.00
70 T	Styrene	50.000	52.083	-4.2	86	0.00
71 P	Bromoform	50.000	57.709	-15.4	92	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	86	0.00
73 T	Isopropylbenzene	50.000	47.871	4.3	82	0.00
74 T	N-amyl acetate	50.000	58.179	-16.4	95	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	55.151	-10.3	96	0.00
76 T	1,2,3-Trichloropropane	50.000	55.711	-11.4	94	0.00
77 T	Bromobenzene	50.000	51.070	-2.1	87	0.00
78 T	n-propylbenzene	50.000	48.235	3.5	81	0.00
79 T	2-Chlorotoluene	50.000	48.165	3.7	83	0.00
80 T	1,3,5-Trimethylbenzene	50.000	48.561	2.9	82	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.891	0.2	78	0.00
82 T	4-Chlorotoluene	50.000	47.930	4.1	83	0.00
83 T	tert-Butylbenzene	50.000	48.556	2.9	83	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.335	1.3	84	0.00
85 T	sec-Butylbenzene	50.000	48.219	3.6	81	0.00
86 T	p-Isopropyltoluene	50.000	48.753	2.5	81	0.00
87 T	1,3-Dichlorobenzene	50.000	49.770	0.5	84	0.00
88 T	1,4-Dichlorobenzene	50.000	50.109	-0.2	84	0.00
89 T	n-Butylbenzene	50.000	48.368	3.3	80	0.00
90 T	Hexachloroethane	50.000	48.300	3.4	76	0.00
91 T	1,2-Dichlorobenzene	50.000	50.635	-1.3	86	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	53.822	-7.6	92	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.606	-1.2	83	0.00
94 T	Hexachlorobutadiene	50.000	46.603	6.8	77	0.00
95 T	Naphthalene	50.000	55.882	-11.8	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061822\
Data File : VX029586.D
Acq On : 18 Jun 2022 21:39
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 20 01:37:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	51.090	-2.2	85	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6