

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX061822\  
 Data File : VX029577.D  
 Acq On : 18 Jun 2022 18:09  
 Operator : JC/MD  
 Sample : N3290-02  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 25 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 MW-2

Manual Integrations  
 APPROVED

Reviewed By : Semsettin Yesilyurt 06/23/2022  
 Supervised By : Mahesh Dadoda 06/23/2022

Quant Time: Jun 20 01:35:33 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X061822W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Jun 20 01:31:01 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 5.556  | 168   | 266214   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 445537   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 393654   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 185795   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 153970   | 43.692   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =      | 87.380%  |
| 35) Dibromofluoromethane    | 5.391  | 113   | 134609   | 46.361   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =      | 92.720%  |
| 50) Toluene-d8              | 8.653  | 98    | 512765   | 48.342   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =      | 96.680%  |
| 62) 4-Bromofluorobenzene    | 11.085 | 95    | 193798   | 48.449   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =      | 96.900%  |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          |        | Qvalue   |
| 11) Tert butyl alcohol      | 2.965  | 59    | 1284240  | 1726.985 | ug/l   | 99       |
| 16) Acetone                 | 2.386  | 43    | 16326    | 11.623   | ug/l   | 100      |
| 17) Carbon Disulfide        | 2.520  | 76    | 2139     | 0.306    | ug/l # | 92       |
| 19) Methyl tert-butyl Ether | 3.117  | 73    | 233585   | 24.947   | ug/l # | 49       |
| 22) Diisopropyl ether       | 3.763  | 45    | 47307    | 5.492    | ug/l # | 87       |
| 25) 2-Butanone              | 4.580  | 43    | 12475    | 5.619    | ug/l   | 91       |
| 31) Cyclohexane             | 5.477  | 56    | 66675    | 14.093   | ug/l # | 96       |
| 39) Methylcyclohexane       | 7.385  | 83    | 64722    | 13.210   | ug/l   | 92       |
| 40) Benzene                 | 6.044  | 78    | 2441000  | 209.553  | ug/l   | 99       |
| 52) Toluene                 | 8.720  | 92    | 135101   | 18.112   | ug/l   | 95       |
| 67) Ethyl Benzene           | 10.195 | 91    | 1584452  | 114.029  | ug/l   | 97       |
| 68) m/p-Xylenes             | 10.305 | 106   | 385520   | 71.145   | ug/l   | 97       |
| 69) o-Xylene                | 10.646 | 106   | 163449   | 30.530   | ug/l   | 97       |
| 73) Isopropylbenzene        | 10.963 | 105   | 288609   | 20.718   | ug/l   | 98       |
| 78) n-propylbenzene         | 11.305 | 91    | 845229   | 53.871   | ug/l   | 100      |
| 80) 1,3,5-Trimethylbenzene  | 11.451 | 105   | 409958   | 34.944   | ug/l   | 99       |
| 84) 1,2,4-Trimethylbenzene  | 11.756 | 105   | 618665   | 52.874   | ug/l   | 100      |
| 85) sec-Butylbenzene        | 11.890 | 105   | 45314    | 3.196    | ug/l   | 97       |
| 86) p-Isopropyltoluene      | 12.012 | 119   | 15517m   | 1.307    | ug/l   |          |
| 89) n-Butylbenzene          | 12.329 | 91    | 75053    | 7.492    | ug/l # | 36       |
| 95) Naphthalene             | 13.780 | 128   | 636312   | 48.825   | ug/l   | 99       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX061822\  
Data File : VX029577.D  
Acq On : 18 Jun 2022 18:09  
Operator : JC/MD  
Sample : N3290-02  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 25 Sample Multiplier: 1


Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-2

Quant Time: Jun 20 01:35:33 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X061822W.M  
Quant Title : SW846 8260  
QLast Update : Mon Jun 20 01:31:01 2022  
Response via : Initial Calibration

Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 06/23/2022  
Supervised By :Mahesh Dadoda 06/23/2022





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.556 min Scan# 714  
Delta R.T. -0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09



Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-2



Tgt Ion:168 Resp: 266214  
Ion Ratio Lower Upper  
168 100  
99 62.1 41.7 62.5


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/23/2022  
Supervised By :Mahesh Dadoda 06/23/2022






06/23/2022  
Supervised By :Mahesh  
Dadoda


06/23/2022



#11  
Tert butyl alcohol  
Concen: 1726.985 ug/l  
RT: 2.965 min Scan# 308  
Delta R.T. -0.018 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09


Tgt Ion: 59 Resp: 1284240  
Ion Ratio Lower Upper  
59 100  
57 10.1 8.3 12.5





#16  
Acetone  
Concen: 11.623 ug/l  
RT: 2.386 min Scan# 213  
Delta R.T. -0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09



Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-2



Tgt Ion: 43 Resp: 16320  
Ion Ratio Lower Upper  
43 100  
58 36.1 28.8 43.2


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/23/2022  
Supervised By :Mahesh Dadoda 06/23/2022




06/23/2022  
Supervised By :Mahesh  
Dadoda



06/23/2022




#17  
Carbon Disulfide  
Concen: 0.306 ug/l  
RT: 2.520 min Scan# 235  
Delta R.T. 0.006 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09




Tgt Ion: 76 Resp: 2139  
Ion Ratio Lower Upper  
76 100  
78 12.1 7.2 10.8#





#19  
Methyl tert-butyl Ether  
Concen: 24.947 ug/l  
RT: 3.117 min Scan# 311  
Delta R.T. 0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09



Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-2



Tgt Ion: 73 Resp: 23358  
Ion Ratio Lower Upper  
73 100  
57 45.0 16.8 25.2


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/23/2022  
Supervised By :Mahesh Dadoda 06/23/2022




06/23/2022  
Supervised By :Mahesh  
Dadoda



06/23/2022




#22  
Diisopropyl ether  
Concen: 5.492 ug/l  
RT: 3.763 min Scan# 439  
Delta R.T. 0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09




Tgt Ion: 45 Resp: 47307  
Ion Ratio Lower Upper  
45 100  
43 48.0 50.8 76.2#  
87 30.8 26.6 39.8  
59 14.3 10.5 15.7





#25  
2-Butanone  
Concen: 5.619 ug/l  
RT: 4.580 min Scan# 51  
Delta R.T. 0.018 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09



Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-2



Tgt Ion: 43 Resp: 12475  
Ion Ratio Lower Upper  
43 100  
72 25.7 24.4 36.6


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/23/2022  
Supervised By :Mahesh Dadoda 06/23/2022






06/23/2022  
Supervised By :Mahesh  
Dadoda


06/23/2022



#31  
Cyclohexane  
Concen: 14.093 ug/l  
RT: 5.477 min Scan# 720  
Delta R.T. -0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09

Tgt Ion: 56 Resp: 66675  
Ion Ratio Lower Upper  
56 100  
69 25.1 26.7 40.1#  
84 97.3 77.2 115.8





#33

1,2-Dichloroethane-d4

Concen: 43.692 ug/l

RT: 5.958 min Scan# 799

Delta R.T. 0.000 min

Lab File: VX029577.D

Acq: 18 Jun 2022 18:09

Instrument :

MSVOA\_X

ClientSampleId :

MW-2

Tgt Ion: 65 Resp: 153970

Ion Ratio Lower Upper

65 100



67 52.8 0.0 107.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/23/2022



Supervised By :Mahesh Dadoda 06/23/2022



06/23/2022

Supervised By :Mahesh  
Dadoda

06/23/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.000 min

Lab File: VX029577.D

Acq: 18 Jun 2022 18:09




Tgt Ion:114 Resp: 445537


Ion Ratio Lower Upper

114 100

63 19.3 0.0 36.8


88 16.6 0.0 32.4





#35  
Dibromofluoromethane  
Concen: 46.361 ug/l  
RT: 5.391 min Scan# 706  
Delta R.T. 0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09



Instrument :  
MSVOA\_X  
Client Sample Id :  
MW-2



Tgt Ion: 113 Resp: 134609  
Ion Ratio Lower Upper  
113 100  
111 102.1 82.8 124.2  
192 18.2 16.9 25.3


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/23/2022  
Supervised By :Mahesh Dadoda 06/23/2022






06/23/2022  
Supervised By :Mahesh  
Dadoda

06/23/2022




#39  
Methylcyclohexane  
Concen: 13.210 ug/l  
RT: 7.385 min Scan# 1033  
Delta R.T. -0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09

Tgt Ion: 83 Resp: 64722  
Ion Ratio Lower Upper  
83 100  
55 75.0 54.6 81.8  
98 51.8 37.4 56.2









#40  
Benzene  
Concen: 209.553 ug/l  
RT: 6.044 min Scan# 813  
Delta R.T. -0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09

Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-2

Tgt Ion: 78 Resp: 2441000  
Ion Ratio Lower Upper  
78 100  
77 23.8 19.3 28.9



Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/23/2022  
Supervised By :Mahesh Dadoda 06/23/2022






06/23/2022  
Supervised By :Mahesh  
Dadoda


06/23/2022



#50  
Toluene-d8  
Concen: 48.342 ug/l  
RT: 8.653 min Scan# 1241  
Delta R.T. -0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09


Tgt Ion: 98 Resp: 512765  
Ion Ratio Lower Upper  
98 100  
100 66.6 52.6 78.8





#52  
Toluene  
Concen: 18.112 ug/l  
RT: 8.720 min Scan# 11  
Delta R.T. -0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09



Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-2



Tgt Ion: 92 Resp: 13510  
Ion Ratio Lower Upper  
92 100  
91 165.5 137.5 206.3


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/23/2022  
Supervised By :Mahesh Dadoda 06/23/2022






06/23/2022  
Supervised By :Mahesh  
Dadoda


06/23/2022



#62  
4-Bromofluorobenzene  
Concen: 48.449 ug/l  
RT: 11.085 min Scan# 1640  
Delta R.T. 0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09


Tgt Ion: 95 Resp: 193798  
Ion Ratio Lower Upper  
95 100  
174 70.2 0.0 163.0  
176 68.2 0.0 155.0





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. -0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09



Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-2



Tgt Ion:117 Resp: 393654  
Ion Ratio Lower Upper  
117 100  
82 56.0 41.8 62.8  
119 32.1 24.8 37.2


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/23/2022  
Supervised By :Mahesh Dadoda 06/23/2022






06/23/2022  
Supervised By :Mahesh  
Dadoda


06/23/2022



#67  
Ethyl Benzene  
Concen: 114.029 ug/l  
RT: 10.195 min Scan# 1494  
Delta R.T. 0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09


Tgt Ion: 91 Resp: 1584452  
Ion Ratio Lower Upper  
91 100  
106 32.2 24.6 36.8





#68  
m/p-Xylenes  
Concen: 71.145 ug/l  
RT: 10.305 min Scan# 1512  
Delta R.T. -0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09



Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-2



Tgt Ion:106 Resp: 385520  
Ion Ratio Lower Upper  
106 100  
91 205.5 160.4 240.6


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/23/2022  
Supervised By :Mahesh Dadoda 06/23/2022






06/23/2022  
Supervised By :Mahesh  
Dadoda


06/23/2022



#69  
o-Xylene  
Concen: 30.530 ug/l  
RT: 10.646 min Scan# 1568  
Delta R.T. 0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09


Tgt Ion:106 Resp: 163449  
Ion Ratio Lower Upper  
106 100  
91 216.8 106.1 318.1





#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.024 min Scan# 1794  
Delta R.T. -0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09



Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-2



Tgt Ion:152 Resp: 185799  
Ion Ratio Lower Upper  
152 100  
115 62.6 37.5 112.7  
150 155.1 0.0 338.2


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/23/2022  
Supervised By :Mahesh Dadoda 06/23/2022




06/23/2022  
Supervised By :Mahesh  
Dadoda



06/23/2022




#73  
Isopropylbenzene  
Concen: 20.718 ug/l  
RT: 10.963 min Scan# 1620  
Delta R.T. 0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09




Tgt Ion:105 Resp: 288609  
Ion Ratio Lower Upper  
105 100  
120 26.2 13.5 40.5





#78  
n-propylbenzene  
Concen: 53.871 ug/l  
RT: 11.305 min Scan# 1676  
Delta R.T. -0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09



Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-2



Tgt Ion: 91 Resp: 845229  
Ion Ratio Lower Upper  
91 100  
120 23.1 11.6 34.8


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/23/2022  
Supervised By :Mahesh Dadoda 06/23/2022




06/23/2022  
Supervised By :Mahesh  
Dadoda



06/23/2022




#80  
1,3,5-Trimethylbenzene  
Concen: 34.944 ug/l  
RT: 11.451 min Scan# 1700  
Delta R.T. 0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09




Tgt Ion:105 Resp: 409958  
Ion Ratio Lower Upper  
105 100  
120 48.0 24.3 72.9





#84  
1,2,4-Trimethylbenzene  
Concen: 52.874 ug/l  
RT: 11.756 min Scan# 1750  
Delta R.T. -0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09



Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-2



Tgt Ion:105 Resp: 61866  
Ion Ratio Lower Upper  
105 100  
120 44.9 22.6 67.8


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/23/2022  
Supervised By :Mahesh Dadoda 06/23/2022






06/23/2022  
Supervised By :Mahesh  
Dadoda


06/23/2022



#85  
sec-Butylbenzene  
Concen: 3.196 ug/l  
RT: 11.890 min Scan# 1772  
Delta R.T. 0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09


Tgt Ion:105 Resp: 45314  
Ion Ratio Lower Upper  
105 100  
134 18.8 10.2 30.4





#86  
p-Isopropyltoluene  
Concen: 1.307 ug/l m  
RT: 12.012 min Scan# 119  
Delta R.T. -0.000 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09



Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-2



Tgt Ion:119 Resp: 1551  
Ion Ratio Lower Upper  
119 100  
134 78.0 13.0 39.0  
91 68.7 11.3 33.9


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/23/2022  
Supervised By :Mahesh Dadoda 06/23/2022




06/23/2022  
Supervised By :Mahesh  
Dadoda



06/23/2022




#89  
n-Butylbenzene  
Concen: 7.492 ug/l  
RT: 12.329 min Scan# 1844  
Delta R.T. -0.006 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09



Tgt Ion: 91 Resp: 75053  
Ion Ratio Lower Upper  
91 100  
92 34.8 27.8 83.3  
134 96.9 13.3 39.8








#95  
Naphthalene  
Concen: 48.825 ug/l  
RT: 13.780 min Scan# 2082  
Delta R.T. 0.006 min  
Lab File: VX029577.D  
Acq: 18 Jun 2022 18:09



Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-2



Tgt Ion:128 Resp: 636312  
Ion Ratio Lower Upper  
128 100  
127 13.0 10.2 15.4  
129 11.1 8.7 13.1

Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/23/2022  
Supervised By :Mahesh Dadoda 06/23/2022



06/23/2022  
Supervised By :Mahesh  
Dadoda

06/23/2022