

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092722\
 Data File : VX031694.D
 Acq On : 27 Sep 2022 11:40
 Operator : JC/MD
 Sample : N4790-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 250FERRYBLVD-SOIL-1

Quant Time: Sep 28 06:29:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	50207	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	84230	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	96554	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	56276	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67051	62.923	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 125.840%#	
35) Dibromofluoromethane	5.379	113	52336	52.147	ug/l	-0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.300%	
50) Toluene-d8	8.647	98	183114	50.326	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.660%	
62) 4-Bromofluorobenzene	11.079	95	63324	47.484	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.960%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	37010	141.771	ug/l	95
18) Methyl Acetate	2.703	43	10516	10.982	ug/l	92
20) Methylene Chloride	2.788	84	11726	16.259	ug/l	94
25) 2-Butanone	4.562	43	9276	18.237	ug/l	96
37) Ethyl Acetate	4.715	43	10288	8.849	ug/l	96
43) Isopropyl Acetate	6.342	43	47364	25.885	ug/l	97
52) Toluene	8.720	92	8799	3.729	ug/l	92
65) Chlorobenzene	10.079	112	2949	1.096	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092722\
Data File : VX031694.D
Acq On : 27 Sep 2022 11:40
Operator : JC/MD
Sample : N4790-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
250FERRYBLVD-SOIL-1

Quant Time: Sep 28 06:29:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

