

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092722\
 Data File : VX031695.D
 Acq On : 27 Sep 2022 12:03
 Operator : JC/MD
 Sample : N4809-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP1-0923

Quant Time: Sep 28 06:29:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	60665	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	96597	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	115962	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68943	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69012	53.599	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.200%
35) Dibromofluoromethane	5.385	113	55252	48.005	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.000%
50) Toluene-d8	8.647	98	211043	50.576	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.160%
62) 4-Bromofluorobenzene	11.079	95	74421	48.661	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.320%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	7129	5.306	ug/l	99
20) Methylene Chloride	2.788	84	7074	7.241	ug/l	88
37) Ethyl Acetate	4.721	43	10062	7.547	ug/l	98
43) Isopropyl Acetate	6.342	43	43355	20.661	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092722\
Data File : VX031695.D
Acq On : 27 Sep 2022 12:03
Operator : JC/MD
Sample : N4809-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP1-0923

Quant Time: Sep 28 06:29:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

