

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092722\
 Data File : VX031696.D
 Acq On : 27 Sep 2022 12:27
 Operator : JC/MD
 Sample : N4809-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP2-0923

Quant Time: Sep 28 06:30:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	57758	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	95382	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	113676	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71068	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	67368	54.956	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.920%
35) Dibromofluoromethane	5.385	113	53663	47.218	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.440%
50) Toluene-d8	8.647	98	208988	50.722	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.440%
62) 4-Bromofluorobenzene	11.079	95	71844	47.574	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.140%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	33028	105.363	ug/l	94
20) Methylene Chloride	2.788	84	7183	7.839	ug/l	91
25) 2-Butanone	4.574	43	3499	5.980	ug/l	95
37) Ethyl Acetate	4.721	43	7601	5.774	ug/l #	93
43) Isopropyl Acetate	6.348	43	35412	17.091	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092722\
Data File : VX031696.D
Acq On : 27 Sep 2022 12:27
Operator : JC/MD
Sample : N4809-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP2-0923

Quant Time: Sep 28 06:30:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

