

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092722\
 Data File : VX031697.D
 Acq On : 27 Sep 2022 12:50
 Operator : JC/MD
 Sample : N4809-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP3-0923

Quant Time: Sep 28 06:30:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	53395	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	87853	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	105668	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63202	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	64604	57.007	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.020%
35) Dibromofluoromethane	5.385	113	51338	49.043	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.080%
50) Toluene-d8	8.647	98	187299	49.354	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.700%
62) 4-Bromofluorobenzene	11.079	95	65650	47.198	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.400%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	7050	8.504	ug/l	99
18) Methyl Acetate	2.703	43	1070	1.051	ug/l #	86
20) Methylene Chloride	2.782	84	6681	7.897	ug/l	88
37) Ethyl Acetate	4.715	43	10079	8.312	ug/l	97
43) Isopropyl Acetate	6.342	43	46581	24.408	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092722\
Data File : VX031697.D
Acq On : 27 Sep 2022 12:50
Operator : JC/MD
Sample : N4809-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP3-0923

Quant Time: Sep 28 06:30:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

