

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092722\
 Data File : VX031698.D
 Acq On : 27 Sep 2022 13:39
 Operator : JC/MD
 Sample : N4791-01RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-TB-20220919

Quant Time: Sep 27 16:31:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	50753	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	78187	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	87114	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55940	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	58037	53.878	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.760%			
35) Dibromofluoromethane	5.385	113	46568	49.986	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.980%			
50) Toluene-d8	8.647	98	171340	50.730	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.460%			
62) 4-Bromofluorobenzene	11.079	95	57595	46.526	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 93.060%			
Target Compounds						
30) Chloroform	5.086	83	2636	1.627	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092722\
Data File : VX031698.D
Acq On : 27 Sep 2022 13:39
Operator : JC/MD
Sample : N4791-01RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-TB-20220919

Quant Time: Sep 27 16:31:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

