

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092722\
 Data File : VX031699.D
 Acq On : 27 Sep 2022 14:02
 Operator : JC/MD
 Sample : N4791-08RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-EB-20220922RE

Quant Time: Sep 27 16:45:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	45836	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	73380	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	79591	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	50235	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	55927	57.489	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.980%
35) Dibromofluoromethane	5.385	113	45183	51.677	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.360%
50) Toluene-d8	8.647	98	165166	52.106	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.220%
62) 4-Bromofluorobenzene	11.079	95	54371	46.799	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.600%
Target Compounds						
30) Chloroform	5.092	83	2692	1.840	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092722\
Data File : VX031699.D
Acq On : 27 Sep 2022 14:02
Operator : JC/MD
Sample : N4791-08RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-EB-20220922RE

Quant Time: Sep 27 16:45:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

