

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092722\
 Data File : VX031700.D
 Acq On : 27 Sep 2022 14:26
 Operator : JC/MD
 Sample : N4791-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-GW-503-505

Quant Time: Sep 27 17:08:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	46033	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	70443	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	79000	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	50127	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	45242	46.307	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.620%
35) Dibromofluoromethane	5.385	113	36434	43.408	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.820%
50) Toluene-d8	8.647	98	128037	42.076	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	84.160%#
62) 4-Bromofluorobenzene	11.079	95	42104	37.751	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	75.500%#
Target Compounds						
2) Dichlorodifluoromethane	1.160	85	577	1.030	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092722\
Data File : VX031700.D
Acq On : 27 Sep 2022 14:26
Operator : JC/MD
Sample : N4791-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-GW-503-505

Quant Time: Sep 27 17:08:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

