

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092722\
 Data File : VX031705.D
 Acq On : 27 Sep 2022 16:34
 Operator : JC/MD
 Sample : N4791-06DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-GW-568-570DL

Quant Time: Sep 27 18:21:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	62040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	96903	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	106124	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64591	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	70903	53.847	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.700%	
35) Dibromofluoromethane	5.391	113	55750	48.284	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.560%	
50) Toluene-d8	8.653	98	203593	48.637	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 97.280%	
62) 4-Bromofluorobenzene	11.079	95	67934	44.279	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 88.560%	
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.319	101	1905	2.310	ug/l	Qvalue 89
44) Trichloroethene	7.123	130	37729	32.799	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092722\
Data File : VX031705.D
Acq On : 27 Sep 2022 16:34
Operator : JC/MD
Sample : N4791-06DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-GW-568-570DL

Quant Time: Sep 27 18:21:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

