

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092722\
 Data File : VX031708.D
 Acq On : 27 Sep 2022 17:44
 Operator : JC/MD
 Sample : N4837-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 251563-3

Quant Time: Sep 28 06:31:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	55083	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	89034	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	97349	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	61378	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	67993	58.159	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.320%
35) Dibromofluoromethane	5.385	113	52943	49.906	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.820%
50) Toluene-d8	8.653	98	195489	50.828	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.660%
62) 4-Bromofluorobenzene	11.079	95	65662	46.580	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092722\
Data File : VX031708.D
Acq On : 27 Sep 2022 17:44
Operator : JC/MD
Sample : N4837-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
251563-3

Quant Time: Sep 28 06:31:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

