

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092722\
 Data File : VX031710.D
 Acq On : 27 Sep 2022 18:30
 Operator : JC/MD
 Sample : N4841-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 252672-5

Quant Time: Sep 28 06:32:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	53400	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	86613	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	96415	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60172	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	64769	57.147	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.300%
35) Dibromofluoromethane	5.385	113	50884	49.306	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.620%
50) Toluene-d8	8.653	98	187064	49.997	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.000%
62) 4-Bromofluorobenzene	11.079	95	63211	46.095	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092722\
Data File : VX031710.D
Acq On : 27 Sep 2022 18:30
Operator : JC/MD
Sample : N4841-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
252672-5

Quant Time: Sep 28 06:32:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

