

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122623\
 Data File : VX039466.D
 Acq On : 26 Dec 2023 14:42
 Operator : JC/MD
 Sample : 06014-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AUD-1306

Quant Time: Dec 27 04:38:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	107539	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	223619	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226554	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102566	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	94819	48.131	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.260%
35) Dibromofluoromethane	5.379	113	72934	47.211	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.420%
50) Toluene-d8	8.647	98	286184	49.898	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.800%
62) 4-Bromofluorobenzene	11.079	95	121165	51.202	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.400%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.989	59	36685	74.155	ug/l #	92
17) Carbon Disulfide	2.508	76	2608	0.644	ug/l #	93
18) Methyl Acetate	2.709	43	1477	0.347	ug/l	96
19) Methyl tert-butyl Ether	3.117	73	2933	0.586	ug/l #	77
20) Methylene Chloride	2.794	84	1110	0.676	ug/l	92
40) Benzene	6.037	78	185593	26.913	ug/l	99
52) Toluene	8.720	92	5150	1.237	ug/l	96
67) Ethyl Benzene	10.195	91	3104	0.337	ug/l #	83
68) m/p-Xylenes	10.305	106	12511	3.791	ug/l	89
69) o-Xylene	10.640	106	2303	0.722	ug/l	95
73) Isopropylbenzene	10.963	105	3753	0.474	ug/l	98
78) n-propylbenzene	11.311	91	4813	0.494	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	7890	1.170	ug/l	94
84) 1,2,4-Trimethylbenzene	11.756	105	17907	2.590	ug/l	99
95) Naphthalene	13.774	128	27936	3.724	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122623\
Data File : VX039466.D
Acq On : 26 Dec 2023 14:42
Operator : JC/MD
Sample : 06014-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AUD-1306

Quant Time: Dec 27 04:38:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

