

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122623\
 Data File : VX039471.D
 Acq On : 26 Dec 2023 16:35
 Operator : JC/MD
 Sample : 06019-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 27 04:40:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	111034	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	233949	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	231821	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101315	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	98876	48.610	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.220%
35) Dibromofluoromethane	5.385	113	74986	46.396	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.800%
50) Toluene-d8	8.647	98	298254	49.706	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.420%
62) 4-Bromofluorobenzene	11.079	95	121432	49.049	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122623\
Data File : VX039471.D
Acq On : 26 Dec 2023 16:35
Operator : JC/MD
Sample : 06019-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 27 04:40:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

