

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122623\
 Data File : VX039474.D
 Acq On : 26 Dec 2023 17:44
 Operator : JC/MD
 Sample : PB158014TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB158014TB

Quant Time: Dec 27 04:41:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	89913	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	189415	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	194272	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83211	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78253	47.508	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.020%
35) Dibromofluoromethane	5.385	113	62049	47.417	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.840%
50) Toluene-d8	8.647	98	245791	50.594	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.180%
62) 4-Bromofluorobenzene	11.079	95	99600	49.690	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.380%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.953	59	3872	9.361	ug/l #	92
16) Acetone	2.380	43	36050	15.766	ug/l	95
20) Methylene Chloride	2.788	84	4213	3.068	ug/l	97
25) 2-Butanone	4.568	43	8008	5.634	ug/l #	90
29) Tetrahydrofuran	5.013	42	5674	7.229	ug/l #	82
37) Ethyl Acetate	4.727	43	3709	1.454	ug/l #	89
43) Isopropyl Acetate	6.336	43	114775	27.666	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122623\
Data File : VX039474.D
Acq On : 26 Dec 2023 17:44
Operator : JC/MD
Sample : PB158014TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB158014TB

Quant Time: Dec 27 04:41:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

