

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122623\
 Data File : VX039475.D
 Acq On : 26 Dec 2023 18:06
 Operator : JC/MD
 Sample : PB158014ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB158014ZHE#01

Quant Time: Dec 27 04:42:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	91226	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	195104	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	205277	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89631	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	82157	49.161	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.320%		
35) Dibromofluoromethane	5.385	113	63591	47.179	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.360%		
50) Toluene-d8	8.646	98	256338	51.226	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	102.460%		
62) 4-Bromofluorobenzene	11.079	95	107216	51.930	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	103.860%		
Target Compounds						
11) Tert butyl alcohol	2.940	59	3659	8.719	ug/l #	80
16) Acetone	2.373	43	31522	10.050	ug/l	96
20) Methylene Chloride	2.794	84	13193	9.470	ug/l	97
37) Ethyl Acetate	4.726	43	2851m	1.085	ug/l	
43) Isopropyl Acetate	6.342	43	125217	29.303	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122623\
Data File : VX039475.D
Acq On : 26 Dec 2023 18:06
Operator : JC/MD
Sample : PB158014ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB158014ZHE#01

Quant Time: Dec 27 04:42:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

