

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122623\
 Data File : VX039478.D
 Acq On : 26 Dec 2023 19:15
 Operator : JC/MD
 Sample : PB158014ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB158014ZHE#04

Quant Time: Dec 27 04:43:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	88167	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	191219	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	203258	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95152	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	80969	50.131	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.260%	
35) Dibromofluoromethane	5.385	113	62510	47.319	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 94.640%	
50) Toluene-d8	8.647	98	252240	51.432	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.860%	
62) 4-Bromofluorobenzene	11.079	95	107776	53.261	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.520%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.947	59	2914	7.185	ug/l	# 86
16) Acetone	2.380	43	31540	11.307	ug/l	94
20) Methylene Chloride	2.794	84	12508	9.290	ug/l	96
43) Isopropyl Acetate	6.336	43	123136	29.402	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122623\
Data File : VX039478.D
Acq On : 26 Dec 2023 19:15
Operator : JC/MD
Sample : PB158014ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB158014ZHE#04

Quant Time: Dec 27 04:43:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

