

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122623\
 Data File : VX039479.D
 Acq On : 26 Dec 2023 19:38
 Operator : JC/MD
 Sample : PB158014ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB158014ZHE#05

Quant Time: Dec 27 04:43:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	81531	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	177329	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	188252	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	87617	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	74464	49.856	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.720%
35) Dibromofluoromethane	5.385	113	58695	47.911	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.820%
50) Toluene-d8	8.647	98	232802	51.186	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.380%
62) 4-Bromofluorobenzene	11.079	95	99091	52.805	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.620%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	31053	13.695	ug/l #	88
20) Methylene Chloride	2.788	84	12657	10.166	ug/l	92
37) Ethyl Acetate	4.739	43	3093	1.295	ug/l #	90
43) Isopropyl Acetate	6.336	43	119225	30.698	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122623\
Data File : VX039479.D
Acq On : 26 Dec 2023 19:38
Operator : JC/MD
Sample : PB158014ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB158014ZHE#05

Quant Time: Dec 27 04:43:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

