

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010825\
 Data File : VX044616.D
 Acq On : 08 Jan 2025 10:31
 Operator : JC/MD
 Sample : VX0108MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0108MBL01

Quant Time: Jan 09 00:24:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 08 03:57:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	179403	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	331125	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	312334	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	140885	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	152204	52.617	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	105.240%	
35) Dibromofluoromethane	5.373	113	112676	49.625	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	99.240%	
50) Toluene-d8	8.641	98	407495	52.541	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	105.080%	
62) 4-Bromofluorobenzene	11.079	95	153494	56.071	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	112.140%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010825\
Data File : VX044616.D
Acq On : 08 Jan 2025 10:31
Operator : JC/MD
Sample : VX0108MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0108MBL01

Quant Time: Jan 09 00:24:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010725W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 08 03:57:12 2025
Response via : Initial Calibration

