

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010825\
 Data File : VX044620.D
 Acq On : 08 Jan 2025 12:10
 Operator : JC/MD
 Sample : Q1007-01
 Misc : 1.25g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WD-01

Quant Time: Jan 08 12:39:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 08 03:57:12 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	153959	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.751	114	282200	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	256515	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	134344	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	131096	52.810	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.620%
35) Dibromofluoromethane	5.385	113	88577	45.774	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.540%
50) Toluene-d8	8.641	98	345902	52.331	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.660%
62) 4-Bromofluorobenzene	11.079	95	138130	59.206	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	118.420%
Target Compounds						
16) Acetone	2.416	43	10209	13.048	ug/l	Qvalue 95
17) Carbon Disulfide	2.477	76	748	0.210	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010825\
Data File : VX044620.D
Acq On : 08 Jan 2025 12:10
Operator : JC/MD
Sample : Q1007-01
Misc : 1.25g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WD-01

Quant Time: Jan 08 12:39:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010725W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 08 03:57:12 2025
Response via : Initial Calibration

