

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010825\
 Data File : VX044621.D
 Acq On : 08 Jan 2025 12:33
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jan 09 00:27:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 08 03:57:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	176700	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	324894	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	296033	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	129935	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	145140	50.943	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 101.880%			
35) Dibromofluoromethane	5.385	113	102400	45.964	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 91.920%			
50) Toluene-d8	8.647	98	389814	51.225	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.440%			
62) 4-Bromofluorobenzene	11.079	95	138018	51.384	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 102.760%			
Target Compounds						
16) Acetone	2.380	43	1039	1.157	ug/l #	Qvalue 69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010825\
Data File : VX044621.D
Acq On : 08 Jan 2025 12:33
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jan 09 00:27:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010725W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 08 03:57:12 2025
Response via : Initial Calibration

