

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010825\
 Data File : VX044622.D
 Acq On : 08 Jan 2025 12:56
 Operator : JC/MD
 Sample : Q1006-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
ClientSampleId :
 Storage-Blank-SOIL-REF-6

Quant Time: Jan 09 00:27:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 08 03:57:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	178134	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	328051	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	300351	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	133413	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	148247	51.614	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	103.220%	
35) Dibromofluoromethane	5.379	113	106244	47.230	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	94.460%	
50) Toluene-d8	8.647	98	393640	51.230	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	102.460%	
62) 4-Bromofluorobenzene	11.079	95	141818	52.291	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	104.580%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010825\
 Data File : VX044622.D
 Acq On : 08 Jan 2025 12:56
 Operator : JC/MD
 Sample : Q1006-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
ClientSampleId :
 Storage-Blank-SOIL-REF-6

Quant Time: Jan 09 00:27:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 08 03:57:12 2025
 Response via : Initial Calibration

