

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010825\
 Data File : VX044624.D
 Acq On : 08 Jan 2025 13:42
 Operator : JC/MD
 Sample : Q1006-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-WATER-REF-4

Quant Time: Jan 09 00:28:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 08 03:57:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	181030	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	329662	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	320820	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	129243	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	147321	50.472	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 100.940%		
35) Dibromofluoromethane	5.379	113	104004	46.009	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 92.020%		
50) Toluene-d8	8.647	98	422950	54.775	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 109.560%		
62) 4-Bromofluorobenzene	11.079	95	150892	55.365	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 110.720%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010825\
Data File : VX044624.D
Acq On : 08 Jan 2025 13:42
Operator : JC/MD
Sample : Q1006-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-WATER-REF-4

Quant Time: Jan 09 00:28:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010725W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 08 03:57:12 2025
Response via : Initial Calibration

