

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010825\
 Data File : VX044629.D
 Acq On : 08 Jan 2025 15:36
 Operator : JC/MD
 Sample : Q1029-01ME
 Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HR-02-01072025ME

Quant Time: Jan 09 00:31:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 08 03:57:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	162314	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	298311	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	272158	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	124190	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	131624	50.293	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.580%
35) Dibromofluoromethane	5.385	113	85459	41.778	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	83.560%
50) Toluene-d8	8.647	98	351788	50.347	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.700%
62) 4-Bromofluorobenzene	11.079	95	128411	52.068	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.140%
Target Compounds						
8) Diethyl Ether	2.130	74	996	0.862	ug/l #	40
20) Methylene Chloride	2.776	84	612	0.291	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010825\
Data File : VX044629.D
Acq On : 08 Jan 2025 15:36
Operator : JC/MD
Sample : Q1029-01ME
Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HR-02-01072025ME

Quant Time: Jan 09 00:31:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010725W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 08 03:57:12 2025
Response via : Initial Calibration

