

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX010919\
 Data File : VX006963.D
 Acq On : 09 Jan 2019 14:20
 Operator : JC/MD
 Sample : K1098-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S32-0650

Quant Time: Jan 10 01:31:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	659714	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1001018	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	911467	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	438981	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	342226	49.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.28%	
35) Dibromofluoromethane	5.50	113	311985	48.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.34%	
50) Toluene-d8	8.71	98	1181167	50.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.68%	
62) 4-Bromofluorobenzene	11.13	95	409953	50.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.24%	
Target Compounds						
2) Dichlorodifluoromethane	1.20	85	22249	4.107	ug/l	94
4) Vinyl Chloride	1.41	62	59698	8.800	ug/l	100
16) Acetone	2.45	43	28577	8.384	ug/l	96
18) Methyl Acetate	2.78	43	12038	1.220	ug/l	94
21) trans-1,2-Dichloroethene	3.17	96	189355	28.024	ug/l	96
27) cis-1,2-Dichloroethene	4.60	96	177392	23.801	ug/l	93
44) Trichloroethene	7.21	130	21235	2.669	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX010919\
Data File : VX006963.D
Acq On : 09 Jan 2019 14:20
Operator : JC/MD
Sample : K1098-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S32-0650

Quant Time: Jan 10 01:31:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

