

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010919\  
 Data File : VX006965.D  
 Acq On : 09 Jan 2019 15:13  
 Operator : JC/MD  
 Sample : K1082-10  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 SP-1-5

Quant Time: Jan 10 02:18:40 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X010819W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jan 08 14:06:59 2019  
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	640726	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	992055	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	911714	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	446379	50.00	ug/l	0.00

## System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.07	65	342098	51.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.18%	
35) Dibromofluoromethane	5.51	113	305665	48.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.22%	
50) Toluene-d8	8.71	98	1174070	50.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.98%	
62) 4-Bromofluorobenzene	11.13	95	417060	51.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.88%	

## Target Compounds

					Qvalue
8) Diethyl Ether	2.20	74	9880	2.333	ug/l 100
16) Acetone	2.45	43	31099	9.394	ug/l 93
18) Methyl Acetate	2.78	43	18686	1.949	ug/l 96
20) Methylene Chloride	2.86	84	1624966	260.464	ug/l 95
43) Isopropyl Acetate	6.46	43	33543	2.300	ug/l 97
95) Naphthalene	13.83	128	254827	7.923	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010919\  
 Data File : VX006965.D  
 Acq On : 09 Jan 2019 15:13  
 Operator : JC/MD  
 Sample : K1082-10  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampled :  
 SP-1-5

Quant Time: Jan 10 02:18:40 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X010819W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jan 08 14:06:59 2019  
 Response via : Initial Calibration

