

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010919\
 Data File : VX006971.D
 Acq On : 09 Jan 2019 17:53
 Operator : JC/MD
 Sample : K1006-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JC-03-010819-E

Quant Time: Jan 10 02:34:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	633437	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	987159	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	906224	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	439894	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	337015	50.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.82%	
35) Dibromofluoromethane	5.51	113	299396	47.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.72%	
50) Toluene-d8	8.71	98	1168469	50.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.00%	
62) 4-Bromofluorobenzene	11.13	95	409004	50.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.40%	
Target Compounds						
8) Diethyl Ether	2.19	74	16956	4.049	ug/l	97
16) Acetone	2.45	43	50002	15.279	ug/l	100
18) Methyl Acetate	2.78	43	15996	1.688	ug/l	97
20) Methylene Chloride	2.86	84	204642	32.680	ug/l	95
43) Isopropyl Acetate	6.46	43	22396	1.543	ug/l	99
95) Naphthalene	13.83	128	565215	17.834	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX010919\
Data File : VX006971.D
Acq On : 09 Jan 2019 17:53
Operator : JC/MD
Sample : K1006-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JC-03-010819-E

Quant Time: Jan 10 02:34:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

