

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010919\
 Data File : VX006973.D
 Acq On : 09 Jan 2019 18:46
 Operator : JC/MD
 Sample : K1006-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JC-03-010819-G

Quant Time: Jan 10 02:39:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	630130	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	983660	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	902541	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	423727	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	338466	51.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.80%	
35) Dibromofluoromethane	5.51	113	301603	47.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.76%	
50) Toluene-d8	8.71	98	1163883	50.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.96%	
62) 4-Bromofluorobenzene	11.13	95	399128	49.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.30%	
Target Compounds						
16) Acetone	2.45	43	31940	9.811	ug/l	98
18) Methyl Acetate	2.78	43	18307	1.942	ug/l	95
20) Methylene Chloride	2.86	84	218790	35.165	ug/l	95
43) Isopropyl Acetate	6.46	43	25599	1.770	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX010919\
Data File : VX006973.D
Acq On : 09 Jan 2019 18:46
Operator : JC/MD
Sample : K1006-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JC-03-010819-G

Quant Time: Jan 10 02:39:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

