

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX010919\
 Data File : VX006979.D
 Acq On : 09 Jan 2019 21:26
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 10 02:50:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.411	0.443	-7.8	100	0.00
3 P	Chloromethane	0.478	0.483	-1.0	99	0.00
4 C	Vinyl Chloride	0.514	0.514	0.0	99	0.00
5 T	Bromomethane	0.549	0.490	10.7	96	0.00
6 T	Chloroethane	0.359	0.330	8.1	101	0.00
7 T	Trichlorofluoromethane	0.831	0.832	-0.1	103	0.00
8 T	Diethyl Ether	0.331	0.323	2.4	104	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.497	0.466	6.2	101	0.00
10 T	Methyl Iodide	0.671	0.734	-9.4	102	0.00
11 T	Tert butyl alcohol	0.123	0.109	11.4	95	0.00
12 CM	1,1-Dichloroethene	0.460	0.454	1.3#	103	0.00
13 T	Acrolein	0.056	0.050	10.7	92	0.00
14 T	Allyl chloride	0.788	0.759	3.7	98	0.00
15 T	Acrylonitrile	0.283	0.286	-1.1	106	0.00
16 T	Acetone	0.258	0.240	7.0	99	0.00
17 T	Carbon Disulfide	1.183	1.113	5.9	96	0.00
18 T	Methyl Acetate	0.748	0.854	-14.2	123	0.00
19 T	Methyl tert-butyl Ether	1.595	1.584	0.7	103	0.00
20 T	Methylene Chloride	0.548	0.505	7.8	102	0.00
21 T	trans-1,2-Dichloroethene	0.512	0.489	4.5	103	0.00
22 T	Diisopropyl ether	1.481	1.464	1.1	102	0.00
23 T	Vinyl Acetate	1.249	1.222	2.2	100	0.00
24 P	1,1-Dichloroethane	0.828	0.816	1.4	102	0.00
25 T	2-Butanone	0.370	0.362	2.2	100	0.00
26 T	2,2-Dichloropropane	0.637	0.548	14.0	89	0.00
27 T	cis-1,2-Dichloroethene	0.565	0.546	3.4	100	0.00
28 T	Bromochloromethane	0.363	0.361	0.6	95	0.00
29 T	Tetrahydrofuran	0.240	0.239	0.4	103	0.00
30 C	Chloroform	0.867	0.870	-0.3#	105	0.00
31 T	Cyclohexane	0.737	0.724	1.8	101	0.00
32 T	1,1,1-Trichloroethane	0.735	0.740	-0.7	102	0.00
33 S	1,2-Dichloroethane-d4	0.523	0.510	2.5	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.320	0.312	2.5	92	0.00
36 T	1,1-Dichloropropene	0.435	0.413	5.1	99	0.00
37 T	Ethyl Acetate	0.466	0.463	0.6	102	0.00
38 T	Carbon Tetrachloride	0.437	0.431	1.4	102	0.00
39 T	Methylcyclohexane	0.554	0.516	6.9	99	0.00
40 TM	Benzene	1.326	1.287	2.9	102	0.00
41 T	Methacrylonitrile	0.268	0.256	4.5	105	0.00
42 TM	1,2-Dichloroethane	0.458	0.439	4.1	102	0.00
43 T	Isopropyl Acetate	0.735	0.740	-0.7	102	0.00
44 TM	Trichloroethene	0.397	0.374	5.8	100	0.00
45 C	1,2-Dichloropropane	0.333	0.331	0.6#	103	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX010919\
 Data File : VX006979.D
 Acq On : 09 Jan 2019 21:26
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 10 02:50:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.234	0.232	0.9	103	0.00
47 T	Bromodichloromethane	0.399	0.412	-3.3	104	0.00
48 T	Methyl methacrylate	0.378	0.380	-0.5	104	0.00
49 T	1,4-Dioxane	0.010	0.007	30.0#	86	0.00
50 S	Toluene-d8	1.172	1.140	2.7	91	0.00
51 T	4-Methyl-2-Pentanone	0.472	0.473	-0.2	103	0.00
52 CM	Toluene	0.853	0.826	3.2#	102	0.00
53 T	t-1,3-Dichloropropene	0.442	0.449	-1.6	99	0.00
54 T	cis-1,3-Dichloropropene	0.489	0.501	-2.5	102	0.00
55 T	1,1,2-Trichloroethane	0.341	0.346	-1.5	103	0.00
56 T	Ethyl methacrylate	0.498	0.505	-1.4	102	0.00
57 T	1,3-Dichloropropane	0.559	0.554	0.9	102	0.00
58 T	2-Chloroethyl Vinyl ether	0.256	0.253	1.2	94	0.00
59 T	2-Hexanone	0.360	0.354	1.7	102	0.00
60 T	Dibromochloromethane	0.331	0.360	-8.8	104	0.00
61 T	1,2-Dibromoethane	0.376	0.376	0.0	102	0.00
62 S	4-Bromofluorobenzene	0.409	0.396	3.2	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.433	0.401	7.4	101	0.00
65 PM	Chlorobenzene	1.091	1.051	3.7	102	0.00
66 T	1,1,1,2-Tetrachloroethane	0.367	0.380	-3.5	102	0.00
67 C	Ethyl Benzene	1.800	1.759	2.3#	103	0.00
68 T	m/p-Xylenes	0.709	0.687	3.1	102	0.00
69 T	o-Xylene	0.692	0.671	3.0	102	0.00
70 T	Styrene	1.096	1.089	0.6	103	0.00
71 P	Bromoform	0.281	0.290	-3.2	101	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.526	3.512	0.4	102	0.00
74 T	N-amyl acetate	1.420	1.431	-0.8	101	0.00
75 P	1,1,2,2-Tetrachloroethane	1.119	1.106	1.2	103	0.00
76 T	1,2,3-Trichloropropane	1.013	1.033	-2.0	103	0.00
77 T	Bromobenzene	0.971	0.947	2.5	100	0.00
78 T	n-propylbenzene	3.870	3.916	-1.2	102	0.00
79 T	2-Chlorotoluene	2.330	2.268	2.7	102	0.00
80 T	1,3,5-Trimethylbenzene	2.946	2.913	1.1	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.285	0.295	-3.5	94	0.00
82 T	4-Chlorotoluene	2.698	2.663	1.3	101	0.00
83 T	tert-Butylbenzene	2.939	2.965	-0.9	102	0.00
84 T	1,2,4-Trimethylbenzene	2.962	2.965	-0.1	101	0.00
85 T	sec-Butylbenzene	3.510	3.492	0.5	102	0.00
86 T	p-Isopropyltoluene	3.140	3.124	0.5	102	0.00
87 T	1,3-Dichlorobenzene	1.756	1.680	4.3	101	0.00
88 T	1,4-Dichlorobenzene	1.806	1.684	6.8	103	0.00
89 T	n-Butylbenzene	2.677	2.652	0.9	101	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010919\
Data File : VX006979.D
Acq On : 09 Jan 2019 21:26
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 10 02:50:05 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.443	0.484	-9.3	102	0.00
91 T	1,2-Dichlorobenzene	1.733	1.673	3.5	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.221	0.232	-5.0	102	0.00
93 T	1,2,4-Trichlorobenzene	1.189	1.158	2.6	101	0.00
94 T	Hexachlorobutadiene	0.528	0.514	2.7	101	0.00
95 T	Naphthalene	3.602	3.620	-0.5	104	0.00
96 T	1,2,3-Trichlorobenzene	1.181	1.146	3.0	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5