

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020524\
 Data File : VX040110.D
 Acq On : 05 Feb 2024 13:00
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Feb 06 02:59:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	99056	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	176973	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	169230	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82175	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	70326	44.114	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.220%
35) Dibromofluoromethane	5.385	113	55369	45.960	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.920%
50) Toluene-d8	8.646	98	214918	47.703	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.400%
62) 4-Bromofluorobenzene	11.079	95	82862	47.765	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020524\
Data File : VX040110.D
Acq On : 05 Feb 2024 13:00
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Feb 06 02:59:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

