

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020524\
 Data File : VX040113.D
 Acq On : 05 Feb 2024 14:08
 Operator : JC/MD
 Sample : P1141-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-3-0-2

Quant Time: Feb 06 03:00:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	109017	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	193794	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	180071	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83512	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	76842	43.797	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.600%
35) Dibromofluoromethane	5.385	113	61124	46.333	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.660%
50) Toluene-d8	8.647	98	229381	46.494	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.980%
62) 4-Bromofluorobenzene	11.079	95	88917	46.807	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.620%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	16276	19.799	ug/l	97
20) Methylene Chloride	2.794	84	4276	2.927	ug/l	89
37) Ethyl Acetate	4.739	43	2697	1.294	ug/l #	76
43) Isopropyl Acetate	6.342	43	69581	21.176	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020524\
Data File : VX040113.D
Acq On : 05 Feb 2024 14:08
Operator : JC/MD
Sample : P1141-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-3-0-2

Quant Time: Feb 06 03:00:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

