

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020524\
 Data File : VX040118.D
 Acq On : 05 Feb 2024 16:03
 Operator : JC/MD
 Sample : P1141-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-2-0-2D

Quant Time: Feb 06 03:02:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	91826	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	169338	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	101740	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	48225	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	64772	43.829	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	87.660%		
35) Dibromofluoromethane	5.385	113	50719	43.999	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	88.000%		
50) Toluene-d8	8.653	98	165368	38.360	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	76.720%#		
62) 4-Bromofluorobenzene	11.079	95	77730	46.827	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	93.660%		
Target Compounds						
16) Acetone	2.379	43	12799	18.484	ug/l	96
20) Methylene Chloride	2.788	84	4079	3.315	ug/l	96
37) Ethyl Acetate	4.733	43	3259	1.790	ug/l #	88
43) Isopropyl Acetate	6.342	43	82724	28.811	ug/l	98
52) Toluene	8.720	92	25458	8.412	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020524\
Data File : VX040118.D
Acq On : 05 Feb 2024 16:03
Operator : JC/MD
Sample : P1141-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-2-0-2D

Quant Time: Feb 06 03:02:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

