

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020524\
 Data File : VX040119.D
 Acq On : 05 Feb 2024 16:26
 Operator : JC/MD
 Sample : P1318-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CO-11-GRID-1

Quant Time: Feb 06 03:03:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	91851	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	161207	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	160299	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46925	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65250	44.141	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.280%
35) Dibromofluoromethane	5.385	113	51449	46.883	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.760%
50) Toluene-d8	8.647	98	207098	50.463	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.920%
62) 4-Bromofluorobenzene	11.079	95	79861	50.538	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.080%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	31406	45.344	ug/l	98
20) Methylene Chloride	2.788	84	7991	6.492	ug/l	94
37) Ethyl Acetate	4.727	43	2713	1.565	ug/l #	94
43) Isopropyl Acetate	6.342	43	62342	22.808	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020524\
Data File : VX040119.D
Acq On : 05 Feb 2024 16:26
Operator : JC/MD
Sample : P1318-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CO-11-GRID-1

Quant Time: Feb 06 03:03:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

