

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020524\
 Data File : VX040125.D
 Acq On : 05 Feb 2024 18:44
 Operator : JC/MD
 Sample : P1337-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-5

Quant Time: Feb 06 03:05:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	45547	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	158476	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	164966	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79226	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	65610	89.506	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 179.020%#			
35) Dibromofluoromethane	5.385	113	52724	48.873	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.740%			
50) Toluene-d8	8.647	98	208115	51.585	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 103.160%			
62) 4-Bromofluorobenzene	11.079	95	45123	29.047	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 58.100%#			
Target Compounds						
16) Acetone	2.380	43	33975	98.921	ug/l	94
18) Methyl Acetate	2.709	43	1192	1.465	ug/l #	76
20) Methylene Chloride	2.794	84	5451	8.931	ug/l	96
25) 2-Butanone	4.581	43	3800	7.959	ug/l #	88
37) Ethyl Acetate	4.727	43	3620	2.125	ug/l #	89
43) Isopropyl Acetate	6.342	43	75889	28.243	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020524\
Data File : VX040125.D
Acq On : 05 Feb 2024 18:44
Operator : JC/MD
Sample : P1337-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-5

Quant Time: Feb 06 03:05:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

