Quantitation Report (QT Reviewed)

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020524\
Data File : VX040126.D
Acq On : 05 Feb 2024 19:08
Operator : JC/MD
Sample : P1337-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-6

Manual Integrations APPROVED

Reviewed By :John Carlone 02/06/2024
Supervised By :Mahesh Dadoda 02/06/2024

Quant Time: Feb 06 03:05:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.556	168	87356	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	132708	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	135969	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71048	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	63596	45.236	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 90.480%			
35) Dibromofluoromethane	5.385	113	48441	53.621	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 107.240%			
50) Toluene-d8	8.647	98	123550	36.570	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 73.140%#			
62) 4-Bromofluorobenzene	11.079	95	72536	55.760	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 111.520%			
Target Compounds						Qvalue
16) Acetone	2.380	43	40212	61.045	ug/l #	89
20) Methylene Chloride	2.788	84	8742	7.468	ug/l	94
37) Ethyl Acetate	4.721	43	3316m	2.324	ug/l	
43) Isopropyl Acetate	6.342	43	61912	27.515	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020524\
Data File : VX040126.D
Acq On : 05 Feb 2024 19:08
Operator : JC/MD
Sample : P1337-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Quant Time: Feb 06 03:05:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
TP-6

Manual Integrations APPROVED

Reviewed By :John Carlone 02/06/2024
Supervised By :Mahesh Dadoda 02/06/2024